

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
 Data File : VN072128.D
 Acq On : 21 Apr 2022 16:21
 Operator : JC\MD
 Sample : N2474-02RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TMS-JET-GROUT-SPOILS-ARE

Quant Time: Apr 22 01:24:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/22/2022
 Supervised By : Mahesh Dadoda 04/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	456017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	739192	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	638539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	199618	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	264520	50.811	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.620%			
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount 50.000	Range 69 - 133		Recovery = 0.000%#			
50) Toluene-d8	10.439	98	844425	47.456	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 94.920%			
62) 4-Bromofluorobenzene	12.727	95	236869	41.161	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 82.320%			
Target Compounds					Qvalue	
16) Acetone	4.287	43	282225	126.852	ug/l	89
20) Methylene Chloride	5.098	84	441578	94.165	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
Data File : VN072128.D
Acq On : 21 Apr 2022 16:21
Operator : JC\MD
Sample : N2474-02RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TMS-JET-GROUT-SPOILS-ARE

Quant Time: Apr 22 01:24:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/22/2022
Supervised By :Mahesh Dadoda 04/22/2022

