

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
 Data File : VN072132.D
 Acq On : 21 Apr 2022 17:57
 Operator : JC\MD
 Sample : N2481-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13R

Quant Time: Apr 22 01:25:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	634853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1081269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1095109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	414478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	337805	46.609	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.220%
35) Dibromofluoromethane	8.016	113	307932	48.397	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.800%
50) Toluene-d8	10.439	98	1294371	49.730	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.460%
62) 4-Bromofluorobenzene	12.727	95	447170	53.121	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.240%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	36348	11.735	ug/l	91
25) 2-Butanone	7.351	43	3734	0.836	ug/l #	87
37) Ethyl Acetate	7.416	43	17532	1.803	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
Data File : VN072132.D
Acq On : 21 Apr 2022 17:57
Operator : JC\MD
Sample : N2481-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13R

Quant Time: Apr 22 01:25:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

