

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
 Data File : VN072135.D
 Acq On : 21 Apr 2022 19:09
 Operator : JC\MD
 Sample : N2481-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-26R

Quant Time: Apr 22 01:26:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	592211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1018231	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1098830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	399058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	319516	47.260	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.520%
35) Dibromofluoromethane	8.016	113	291047	48.575	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.140%
50) Toluene-d8	10.439	98	1287354	52.522	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.040%
62) 4-Bromofluorobenzene	12.727	95	458518	57.842	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.680%
Target Compounds						
					Qvalue	
16) Acetone	4.292	43	46503	16.095	ug/l	94
25) 2-Butanone	7.345	43	3849	0.924	ug/l #	73
37) Ethyl Acetate	7.416	43	33964	3.709	ug/l	96
40) Benzene	8.463	78	16259	0.615	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
Data File : VN072135.D
Acq On : 21 Apr 2022 19:09
Operator : JC\MD
Sample : N2481-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-26R

Quant Time: Apr 22 01:26:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

