

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
 Data File : VN072140.D
 Acq On : 21 Apr 2022 21:10
 Operator : JC\MD
 Sample : N2481-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-30

Quant Time: Apr 22 01:27:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	665742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1136248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1195776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	458025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	360797	47.472	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.940%
35) Dibromofluoromethane	8.016	113	324013	48.460	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.920%
50) Toluene-d8	10.439	98	1428155	52.215	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.420%
62) 4-Bromofluorobenzene	12.727	95	501096	56.647	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.300%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	55061	16.952	ug/l	96
25) 2-Butanone	7.345	43	6246	1.334	ug/l	96
37) Ethyl Acetate	7.410	43	40724	3.985	ug/l #	94
95) Naphthalene	15.498	128	16880	4.551	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
Data File : VN072140.D
Acq On : 21 Apr 2022 21:10
Operator : JC\MD
Sample : N2481-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-30

Quant Time: Apr 22 01:27:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

