

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
 Data File : VN072118.D
 Acq On : 21 Apr 2022 11:52
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 22 01:21:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.398	0.506	-27.1#	108	0.00
3 P	Chloromethane	0.504	0.489	3.0	90	0.00
4 C	Vinyl Chloride	0.553	0.513	7.2#	80	0.00
5 T	Bromomethane	0.360	0.337	6.4	81	-0.02
6 T	Chloroethane	0.398	0.360	9.5	80	-0.01
7 T	Trichlorofluoromethane	0.704	0.741	-5.3	92	-0.01
8 T	Diethyl Ether	0.297	0.323	-8.8	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.434	0.481	-10.8	97	0.00
10 T	Methyl Iodide	0.606	0.653	-7.8	91	0.00
11 T	Tert butyl alcohol	0.093	0.088	5.4	78	0.00
12 CM	1,1-Dichloroethene	0.413	0.449	-8.7#	95	0.00
13 T	Acrolein	0.113	0.092	18.6	67	0.00
14 T	Allyl chloride	0.670	0.676	-0.9	86	0.00
15 T	Acrylonitrile	0.269	0.281	-4.5	86	0.00
16 T	Acetone	0.244	0.210	13.9	80	0.00
17 T	Carbon Disulfide	0.947	1.087	-14.8	104	0.00
18 T	Methyl Acetate	0.737	0.628	14.8	89	0.00
19 T	Methyl tert-butyl Ether	1.533	1.587	-3.5	87	0.00
20 T	Methylene Chloride	0.514	0.549	-6.8	96	0.00
21 T	trans-1,2-Dichloroethene	0.453	0.484	-6.8	95	0.00
22 T	Diisopropyl ether	1.458	1.520	-4.3	89	0.00
23 T	Vinyl Acetate	1.160	1.268	-9.3	86	0.00
24 P	1,1-Dichloroethane	0.856	0.919	-7.4	92	0.00
25 T	2-Butanone	0.352	0.351	0.3	84	0.00
26 T	2,2-Dichloropropane	0.735	0.732	0.4	84	0.00
27 T	cis-1,2-Dichloroethene	0.556	0.610	-9.7	95	0.00
28 T	Bromochloromethane	0.345	0.348	-0.9	98	0.00
29 T	Tetrahydrofuran	0.228	0.232	-1.8	83	0.00
30 C	Chloroform	0.904	0.950	-5.1#	91	0.00
31 T	Cyclohexane	0.776	0.818	-5.4	97	0.00
32 T	1,1,1-Trichloroethane	0.773	0.815	-5.4	91	0.00
33 S	1,2-Dichloroethane-d4	0.571	0.546	4.4	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.294	0.318	-8.2	90	0.00
36 T	1,1-Dichloropropene	0.411	0.463	-12.7	94	0.00
37 T	Ethyl Acetate	0.450	0.466	-3.6	81	0.00
38 T	Carbon Tetrachloride	0.403	0.451	-11.9	91	0.00
39 T	Methylcyclohexane	0.489	0.596	-21.9	97	0.00
40 TM	Benzene	1.298	1.468	-13.1	95	0.00
41 T	Methacrylonitrile	0.220	0.226	-2.7	82	0.00
42 TM	1,2-Dichloroethane	0.449	0.467	-4.0	87	0.00
43 T	Isopropyl Acetate	0.748	0.750	-0.3	86	0.00
44 TM	Trichloroethene	0.334	0.379	-13.5	93	0.00
45 C	1,2-Dichloropropane	0.324	0.364	-12.3#	93	0.00
46 T	Dibromomethane	0.220	0.247	-12.3	93	0.00
47 T	Bromodichloromethane	0.433	0.493	-13.9	91	0.00
48 T	Methyl methacrylate	0.310	0.343	-10.6	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
 Data File : VN072118.D
 Acq On : 21 Apr 2022 11:52
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 22 01:21:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	87	0.00
50 S	Toluene-d8	1.204	1.246	-3.5	87	0.00
51 T	4-Methyl-2-Pentanone	0.430	0.479	-11.4	85	0.00
52 CM	Toluene	0.820	0.965	-17.7#	95	0.00
53 T	t-1,3-Dichloropropene	0.454	0.516	-13.7	88	0.00
54 T	cis-1,3-Dichloropropene	0.502	0.589	-17.3	91	0.00
55 T	1,1,2-Trichloroethane	0.338	0.374	-10.7	91	0.00
56 T	Ethyl methacrylate	0.482	0.590	-22.4	93	0.00
57 T	1,3-Dichloropropane	0.565	0.616	-9.0	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.111	0.162	-45.9#	98	0.00
59 T	2-Hexanone	0.303	0.329	-8.6	81	0.00
60 T	Dibromochloromethane	0.329	0.386	-17.3	91	0.00
61 T	1,2-Dibromoethane	0.332	0.379	-14.2	91	0.00
62 S	4-Bromofluorobenzene	0.389	0.423	-8.7	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.330	0.364	-10.3	95	0.00
65 PM	Chlorobenzene	1.001	1.118	-11.7	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.407	-12.7	92	0.00
67 C	Ethyl Benzene	1.695	1.968	-16.1#	92	0.00
68 T	m/p-Xylenes	0.651	0.775	-19.0	93	0.00
69 T	o-Xylene	0.646	0.767	-18.7	93	0.00
70 T	Styrene	0.979	1.224	-25.0#	92	0.00
71 P	Bromoform	0.275	0.336	-22.2	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	4.213	4.578	-8.7	92	0.00
74 T	N-amyl acetate	1.207	1.289	-6.8	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.406	1.413	-0.5	88	0.00
76 T	1,2,3-Trichloropropane	1.158	1.011	12.7	76	0.00
77 T	Bromobenzene	1.056	1.107	-4.8	91	0.00
78 T	n-propylbenzene	4.432	4.941	-11.5	90	0.00
79 T	2-Chlorotoluene	2.842	3.039	-6.9	91	0.00
80 T	1,3,5-Trimethylbenzene	3.338	3.718	-11.4	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.374	0.407	-8.8	84	0.00
82 T	4-Chlorotoluene	2.635	2.881	-9.3	91	0.00
83 T	tert-Butylbenzene	3.052	3.345	-9.6	90	0.00
84 T	1,2,4-Trimethylbenzene	3.257	3.636	-11.6	92	0.00
85 T	sec-Butylbenzene	3.956	4.404	-11.3	89	0.00
86 T	p-Isopropyltoluene	3.154	3.696	-17.2	91	0.00
87 T	1,3-Dichlorobenzene	1.683	1.827	-8.6	89	0.00
88 T	1,4-Dichlorobenzene	1.642	1.778	-8.3	90	0.00
89 T	n-Butylbenzene	2.339	2.729	-16.7	89	0.00
90 T	Hexachloroethane	0.577	0.660	-14.4	92	0.00
91 T	1,2-Dichlorobenzene	1.652	1.799	-8.9	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.230	-5.5	82	0.00
93 T	1,2,4-Trichlorobenzene	0.649	0.834	-28.5#	88	0.00
94 T	Hexachlorobutadiene	0.487	0.510	-4.7	89	0.00
95 T	Naphthalene	1.586	2.360	-48.8#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
Data File : VN072118.D
Acq On : 21 Apr 2022 11:52
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 22 01:21:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.604	0.829	-37.3#	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6