


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042125\
Data File : VN086353.D
Acq On : 21 Apr 2025 08:34
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N041125W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Sat Apr 12 01:21:52 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	10954	PASS
75	95	30	60	50.2	28013	PASS
95	95	100	100	100.0	55779	PASS
96	95	5	9	7.0	3905	PASS
173	174	0.00	2	0.9	339	PASS
174	95	50	100	68.7	38333	PASS
175	174	5	9	7.5	2868	PASS
176	174	95	101	96.6	37011	PASS
177	176	5	9	6.0	2237	PASS