

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042220\
 Data File : VN061135.D
 Acq On : 22 Apr 2020 18:24
 Operator : JC/MD
 Sample : L2271-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-2020-04-13

Quant Time: Apr 23 08:06:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 22 15:26:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.40	168	174576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.35	114	284000	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.21	117	267185	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.15	152	125007	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.77	65	106971	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
35) Dibromofluoromethane	7.32	113	80573	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
50) Toluene-d8	9.88	98	343276	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
62) 4-Bromofluorobenzene	12.21	95	132544	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
Target Compounds						
16) Acetone	3.55	43	12466	2.011	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042220\
Data File : VN061135.D
Acq On : 22 Apr 2020 18:24
Operator : JC/MD
Sample : L2271-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-2020-04-13

Quant Time: Apr 23 08:06:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 22 15:26:40 2020
Response via : Initial Calibration

