

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042220\
 Data File : VN061142.D
 Acq On : 22 Apr 2020 21:02
 Operator : JC/MD
 Sample : L2279-07 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OILY-WATER

Quant Time: Apr 23 08:30:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 22 15:26:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.40	168	183344	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.35	114	300567	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.21	117	287238	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.15	152	140701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.77	65	115544	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	7.31	113	84169	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	9.88	98	368984	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
62) 4-Bromofluorobenzene	12.21	95	143313	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
Target Compounds						
16) Acetone	3.55	43	47578	45.991	ug/l	98
20) Methylene Chloride	4.24	84	312515	147.132	ug/l	99
52) Toluene	9.95	92	6657	1.223	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042220\
Data File : VN061142.D
Acq On : 22 Apr 2020 21:02
Operator : JC/MD
Sample : L2279-07 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OILY-WATER

Quant Time: Apr 23 08:30:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 22 15:26:40 2020
Response via : Initial Calibration

