

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042220\
 Data File : VN061143.D
 Acq On : 22 Apr 2020 21:25
 Operator : JC/MD
 Sample : PB128387TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128387TB

Quant Time: Apr 23 08:34:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 22 15:26:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.40	168	180846	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.35	114	298187	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.21	117	291292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.15	152	144179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.77	65	114224	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	7.31	113	81372	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
50) Toluene-d8	9.88	98	370105	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
62) 4-Bromofluorobenzene	12.21	95	145507	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
Target Compounds						
16) Acetone	3.55	43	21189	12.710	ug/l	98
20) Methylene Chloride	4.24	84	3701	1.767	ug/l	94
43) Isopropyl Acetate	7.92	43	55270	11.142	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042220\
Data File : VN061143.D
Acq On : 22 Apr 2020 21:25
Operator : JC/MD
Sample : PB128387TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128387TB

Quant Time: Apr 23 08:34:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 22 15:26:40 2020
Response via : Initial Calibration

