

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042220\
 Data File : VN061144.D
 Acq On : 22 Apr 2020 21:48
 Operator : JC/MD
 Sample : L2331-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-19A

Quant Time: Apr 23 08:39:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 22 15:26:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.40	168	184171	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.35	114	309123	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.21	117	301978	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.15	152	147052	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.77	65	119540	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
35) Dibromofluoromethane	7.31	113	88442	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
50) Toluene-d8	9.88	98	380708	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	12.21	95	150768	53.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.50%	
Target Compounds						
16) Acetone	3.55	43	20476	11.310	ug/l	98
20) Methylene Chloride	4.25	84	3739	1.752	ug/l	92
43) Isopropyl Acetate	7.92	43	50523	9.824	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042220\
Data File : VN061144.D
Acq On : 22 Apr 2020 21:48
Operator : JC/MD
Sample : L2331-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-19A

Quant Time: Apr 23 08:39:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 22 15:26:40 2020
Response via : Initial Calibration

