

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042220\
 Data File : VN061146.D
 Acq On : 22 Apr 2020 22:33
 Operator : JC/MD
 Sample : L2359-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1B-D

Quant Time: Apr 23 08:46:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 22 15:26:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.40	168	183097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.35	114	307470	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.21	117	297958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.15	152	148355	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.77	65	117599	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
35) Dibromofluoromethane	7.31	113	80450	45.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.70%	
50) Toluene-d8	9.88	98	378222	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	12.21	95	148674	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
Target Compounds						
16) Acetone	3.54	43	20566	11.577	ug/l	98
20) Methylene Chloride	4.24	84	3806	1.794	ug/l	97
43) Isopropyl Acetate	7.92	43	54264	10.609	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042220\
Data File : VN061146.D
Acq On : 22 Apr 2020 22:33
Operator : JC/MD
Sample : L2359-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1B-D

Quant Time: Apr 23 08:46:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042220W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 22 15:26:40 2020
Response via : Initial Calibration

