

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
 Data File : VN066709.D
 Acq On : 22 Apr 2021 12:47
 Operator : JC/MD
 Sample : VN0422WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0422WBL02

Quant Time: Apr 23 06:33:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	316243	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	488888	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	429117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	183423	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	184551	47.77	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.54%
35) Dibromofluoromethane	7.513	113	152344	49.76	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.52%
50) Toluene-d8	10.029	98	589073	49.34	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.68%
62) 4-Bromofluorobenzene	12.349	95	182888	44.73	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
Data File : VN066709.D
Acq On : 22 Apr 2021 12:47
Operator : JC/MD
Sample : VN0422WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0422WBL02

Quant Time: Apr 23 06:33:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

