

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
 Data File : VN066718.D
 Acq On : 22 Apr 2021 16:37
 Operator : JC/MD
 Sample : PB135903TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135903TB

Quant Time: Apr 23 06:35:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	317734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	492043	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	453665	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	191009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	186031	47.93	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.86%
35) Dibromofluoromethane	7.513	113	151973	49.32	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.64%
50) Toluene-d8	10.029	98	609324	50.71	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.42%
62) 4-Bromofluorobenzene	12.346	95	198115	48.14	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.28%
Target Compounds						
					Qvalue	
16) Acetone	3.758	43	26482	18.23	ug/l	97
20) Methylene Chloride	4.472	84	16648	3.85	ug/l	99
43) Isopropyl Acetate	8.095	43	32076	4.18	ug/l #	89
45) 1,2-Dichloropropane	9.203	63	2039	0.51	ug/l #	44
95) Naphthalene	15.084	128	19952	4.03	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
Data File : VN066718.D
Acq On : 22 Apr 2021 16:37
Operator : JC/MD
Sample : PB135903TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135903TB

Quant Time: Apr 23 06:35:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

