

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
 Data File : VN066721.D
 Acq On : 22 Apr 2021 17:48
 Operator : JC/MD
 Sample : M2090-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SEDGWICK-COMP-3

Quant Time: Apr 23 06:36:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	317204	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	486696	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	456613	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.288	152	199570	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	184041	47.49	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.98%
35) Dibromofluoromethane	7.511	113	151325	49.65	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.30%
50) Toluene-d8	10.029	98	599906	50.47	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.94%
62) 4-Bromofluorobenzene	12.346	95	200050	49.15	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.30%
Target Compounds						
						Qvalue
16) Acetone	3.753	43	40418	27.86	ug/l	98
20) Methylene Chloride	4.469	84	18341	4.25	ug/l	97
25) 2-Butanone	6.757	43	13200	6.34	ug/l #	86
43) Isopropyl Acetate	8.093	43	24469	3.22	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
Data File : VN066721.D
Acq On : 22 Apr 2021 17:48
Operator : JC/MD
Sample : M2090-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SEDGWICK-COMP-3

Quant Time: Apr 23 06:36:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

