

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
 Data File : VN066722.D
 Acq On : 22 Apr 2021 18:11
 Operator : JC/MD
 Sample : M2090-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SEDGWICK-COMP-4

Quant Time: Apr 23 06:37:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	305636	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	469424	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	432552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	174832	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	176869	47.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.74%
35) Dibromofluoromethane	7.513	113	145555	49.52	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.04%
50) Toluene-d8	10.029	98	582096	50.78	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.56%
62) 4-Bromofluorobenzene	12.346	95	189695	48.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.64%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	40245	28.79	ug/l	95
20) Methylene Chloride	4.469	84	19262	4.63	ug/l	96
25) 2-Butanone	6.760	43	13559	6.76	ug/l	91
43) Isopropyl Acetate	8.095	43	25198	3.44	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
Data File : VN066722.D
Acq On : 22 Apr 2021 18:11
Operator : JC/MD
Sample : M2090-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SEDGWICK-COMP-4

Quant Time: Apr 23 06:37:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

