

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042221\  
 Data File : VN066716.D  
 Acq On : 22 Apr 2021 15:50  
 Operator : JC/MD  
 Sample : VN0422WBS02  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0422WBS02

Manual Integrations  
 APPROVED

MMDadoda  
 4/23/2021 1:15:33 PM

Quant Time: Apr 23 06:34:19 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042121W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Apr 22 08:25:55 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc   | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|--------|--------|----------|
| -----                        |                |      |            |        |        |          |
| Internal Standards           |                |      |            |        |        |          |
| 1) Pentafluorobenzene        | 7.591          | 168  | 289723     | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.516          | 114  | 470009     | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.348         | 117  | 429268     | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.288         | 152  | 192843     | 50.00  | ug/l   | 0.00     |
|                              |                |      |            |        |        |          |
| System Monitoring Compounds  |                |      |            |        |        |          |
| 33) 1,2-Dichloroethane-d4    | 7.950          | 65   | 176273     | 49.80  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = | 99.60% |        |          |
| 35) Dibromofluoromethane     | 7.513          | 113  | 144107     | 48.96  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = | 97.92% |        |          |
| 50) Toluene-d8               | 10.029         | 98   | 565865     | 49.30  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = | 98.60% |        |          |
| 62) 4-Bromofluorobenzene     | 12.346         | 95   | 188657     | 47.99  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = | 95.98% |        |          |
|                              |                |      |            |        |        |          |
| Target Compounds             |                |      |            |        | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.836          | 85   | 58070      | 17.77  | ug/l   | 97       |
| 3) Chloromethane             | 2.037          | 50   | 79587      | 17.06  | ug/l   | 100      |
| 4) Vinyl Chloride            | 2.163          | 62   | 76387      | 18.41  | ug/l   | 100      |
| 5) Bromomethane              | 2.514          | 94   | 32597      | 13.90  | ug/l   | 93       |
| 6) Chloroethane              | 2.659          | 64   | 45473      | 18.67  | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 2.973          | 101  | 82955      | 18.60  | ug/l   | 100      |
| 8) Diethyl Ether             | 3.356          | 74   | 35116      | 18.51  | ug/l   | 88       |
| 9) 1,1,2-Trichlorotrifluo... | 3.697          | 101  | 51093      | 18.26  | ug/l   | 99       |
| 10) Methyl Iodide            | 3.887          | 142  | 43574      | 9.91   | ug/l   | 97       |
| 11) Tert butyl alcohol       | 4.703          | 59   | 45484      | 101.42 | ug/l   | 100      |
| 12) 1,1-Dichloroethene       | 3.673          | 96   | 51340      | 18.18  | ug/l   | 97       |
| 13) Acrolein                 | 3.549          | 56   | 35134      | 70.90  | ug/l   | 95       |
| 14) Allyl chloride           | 4.247          | 41   | 99693      | 18.42  | ug/l   | 92       |
| 15) Acrylonitrile            | 4.896          | 53   | 144271     | 94.37  | ug/l   | 99       |
| 16) Acetone                  | 3.756          | 43   | 112761     | 85.11  | ug/l   | 99       |
| 17) Carbon Disulfide         | 3.984          | 76   | 163248     | 18.00  | ug/l   | 98       |
| 18) Methyl Acetate           | 4.252          | 43   | 66001      | 17.96  | ug/l   | 98       |
| 19) Methyl tert-butyl Ether  | 4.952          | 73   | 173993     | 18.86  | ug/l   | 96       |
| 20) Methylene Chloride       | 4.469          | 84   | 64174      | 16.29  | ug/l   | 96       |
| 21) trans-1,2-Dichloroethene | 4.952          | 96   | 57174      | 18.43  | ug/l   | 95       |
| 22) Diisopropyl ether        | 5.864          | 45   | 204590     | 18.64  | ug/l   | 93       |
| 23) Vinyl Acetate            | 5.805          | 43   | 832359     | 94.07  | ug/l   | 97       |
| 24) 1,1-Dichloroethane       | 5.759          | 63   | 111346     | 18.27  | ug/l   | 99       |
| 25) 2-Butanone               | 6.746          | 43   | 182563     | 96.06  | ug/l   | 93       |
| 26) 2,2-Dichloropropane      | 6.733          | 77   | 88600      | 17.70  | ug/l   | 100      |
| 27) cis-1,2-Dichloroethene   | 6.744          | 96   | 65508      | 18.21  | ug/l   | 94       |
| 28) Bromochloromethane       | 7.111          | 49   | 55560      | 19.38  | ug/l   | 86       |
| 29) Tetrahydrofuran          | 7.130          | 42   | 126457     | 96.56  | ug/l   | 94       |
| 30) Chloroform               | 7.293          | 83   | 106228     | 18.60  | ug/l   | 98       |
| 31) Cyclohexane              | 7.578          | 56   | 103228     | 17.90  | ug/l   | 95       |
| 32) 1,1,1-Trichloroethane    | 7.494          | 97   | 87699      | 18.38  | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 7.714          | 75   | 77788      | 17.80  | ug/l   | 100      |
| 37) Ethyl Acetate            | 6.843          | 43   | 82384      | 17.94  | ug/l   | 98       |
| 38) Carbon Tetrachloride     | 7.701          | 117  | 74816      | 17.99  | ug/l   | 99       |
| 39) Methylcyclohexane        | 9.015          | 83   | 93458      | 18.04  | ug/l   | 95       |
| 40) Benzene                  | 7.967          | 78   | 255856     | 18.19  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042221\  
 Data File : VN066716.D  
 Acq On : 22 Apr 2021 15:50  
 Operator : JC/MD  
 Sample : VN0422WBS02  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0422WBS02

Manual Integrations  
 APPROVED

MMDadoda  
 4/23/2021 1:15:33 PM

Quant Time: Apr 23 06:34:19 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042121W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Apr 22 08:25:55 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 41) Methacrylonitrile         | 7.090  | 41   | 35852    | 17.57  | ug/l # | 100      |
| 42) 1,2-Dichloroethane        | 8.052  | 62   | 79833    | 17.78  | ug/l   | 100      |
| 43) Isopropyl Acetate         | 8.093  | 43   | 136119   | 18.57  | ug/l # | 90       |
| 44) Trichloroethene           | 8.768  | 130  | 62607    | 18.71  | ug/l   | 100      |
| 45) 1,2-Dichloropropane       | 9.053  | 63   | 68876    | 18.01  | ug/l   | 99       |
| 46) Dibromomethane            | 9.141  | 93   | 41187    | 18.45  | ug/l   | 99       |
| 47) Bromodichloromethane      | 9.337  | 83   | 84316    | 17.97  | ug/l   | 100      |
| 48) Methyl methacrylate       | 9.136  | 41   | 59379    | 18.47  | ug/l   | 92       |
| 49) 1,4-Dioxane               | 9.139  | 88   | 17320    | 445.73 | ug/l   | 91       |
| 51) 4-Methyl-2-Pentanone      | 9.922  | 43   | 387893   | 94.60  | ug/l   | 97       |
| 52) Toluene                   | 10.093 | 92   | 153313   | 18.56  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.321 | 75   | 90535    | 18.28  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 9.777  | 75   | 102944   | 18.24  | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane     | 10.501 | 97   | 58947    | 18.03  | ug/l   | 95       |
| 56) Ethyl methacrylate        | 10.372 | 69   | 80109    | 17.62  | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 10.646 | 76   | 101474   | 18.44  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 9.632  | 63   | 139669   | 167.32 | ug/l   | 96       |
| 59) 2-Hexanone                | 10.694 | 43   | 233798   | 87.56  | ug/l   | 98       |
| 60) Dibromochloromethane      | 10.842 | 129  | 63847    | 18.37  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 10.946 | 107  | 59161    | 18.41  | ug/l   | 99       |
| 64) Tetrachloroethene         | 10.571 | 164  | 57129    | 18.85  | ug/l   | 99       |
| 65) Chlorobenzene             | 11.375 | 112  | 159333   | 18.25  | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 11.450 | 131  | 58911    | 18.59  | ug/l   | 99       |
| 67) Ethyl Benzene             | 11.453 | 91   | 277417   | 18.49  | ug/l   | 99       |
| 68) m/p-Xylenes               | 11.563 | 106  | 207881   | 37.17  | ug/l   | 99       |
| 69) o-Xylene                  | 11.890 | 106  | 103247   | 18.74  | ug/l   | 98       |
| 70) Styrene                   | 11.909 | 104  | 158193   | 17.34  | ug/l   | 99       |
| 71) Bromoform                 | 12.073 | 173  | 44003    | 18.31  | ug/l # | 100      |
| 73) Isopropylbenzene          | 12.193 | 105  | 266757   | 19.46  | ug/l   | 100      |
| 74) N-amyl acetate            | 12.035 | 43   | 29616    | 9.08   | ug/l # | 84       |
| 75) 1,1,2,2-Tetrachloroethane | 12.448 | 83   | 85048    | 18.39  | ug/l   | 97       |
| 76) 1,2,3-Trichloropropane    | 12.499 | 75   | 78746m   | 19.08  | ug/l   |          |
| 77) Bromobenzene              | 12.470 | 156  | 67760    | 19.11  | ug/l   | 99       |
| 78) n-propylbenzene           | 12.534 | 91   | 297931   | 19.50  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 12.617 | 91   | 184262   | 18.60  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene    | 12.679 | 105  | 223181   | 19.16  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 12.250 | 75   | 15384    | 15.65  | ug/l   | 92       |
| 82) 4-Chlorotoluene           | 12.719 | 91   | 179276   | 19.08  | ug/l   | 98       |
| 83) tert-Butylbenzene         | 12.936 | 119  | 187055   | 18.03  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 12.984 | 105  | 211949   | 19.02  | ug/l   | 97       |
| 85) sec-Butylbenzene          | 13.116 | 105  | 247624   | 18.99  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.231 | 119  | 215888   | 19.90  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.229 | 146  | 112395   | 18.83  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 13.306 | 146  | 110698   | 17.25  | ug/l   | 97       |
| 89) n-Butylbenzene            | 13.561 | 91   | 173894   | 18.00  | ug/l   | 99       |
| 90) Hexachloroethane          | 13.816 | 117  | 38208    | 15.75  | ug/l   | 90       |
| 91) 1,2-Dichlorobenzene       | 13.599 | 146  | 112896   | 18.52  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.221 | 75   | 14128    | 17.13  | ug/l   | 89       |
| 93) 1,2,4-Trichlorobenzene    | 14.854 | 180  | 66083    | 19.41  | ug/l   | 96       |
| 94) Hexachlorobutadiene       | 14.950 | 225  | 36341    | 19.31  | ug/l   | 100      |
| 95) Naphthalene               | 15.074 | 128  | 169041   | 20.17  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 15.251 | 180  | 62908    | 19.26  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042221\  
Data File : VN066716.D  
Acq On : 22 Apr 2021 15:50  
Operator : JC/MD  
Sample : VN0422WBS02  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02

Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM

Quant Time: Apr 23 06:34:19 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042121W.M  
Quant Title : SW846 8260  
QLast Update : Thu Apr 22 08:25:55 2021  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1  
Pentafluorobenzene  
Concen: 50.00 ug/l  
RT: 7.591 min Scan# 2188  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 168 Resp: 289723  
Ion Ratio Lower Upper  
168 100  
99 50.3 42.6 63.8




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#2  
Dichlorodifluoromethane  
Concen: 17.77 ug/l  
RT: 1.836 min Scan# 42  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 85 Resp: 58070  
Ion Ratio Lower Upper  
85 100  
87 31.5 16.6 49.8









#3  
Chloromethane  
Concen: 17.06 ug/l  
RT: 2.037 min Scan# 117  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50





Tgt Ion: 50 Resp: 79587  
Ion Ratio Lower Upper  
50 100  
52 32.1 25.8 38.8



#4  
Vinyl Chloride  
Concen: 18.41 ug/l  
RT: 2.163 min Scan# 164  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 62 Resp: 76387  
Ion Ratio Lower Upper  
62 100  
64 31.3 25.1 37.7



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02


Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#5  
Bromomethane  
Concen: 13.90 ug/l  
RT: 2.514 min Scan# 295  
Delta R.T. 0.011 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 94 Resp: 32597  
Ion Ratio Lower Upper  
94 100  
96 97.9 73.0 109.4




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#6  
Chloroethane  
Concen: 18.67 ug/l  
RT: 2.659 min Scan# 349  
Delta R.T. 0.005 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 64 Resp: 45473  
Ion Ratio Lower Upper  
64 100  
66 32.1 25.4 38.2









#7  
Trichlorofluoromethane  
Concen: 18.60 ug/l  
RT: 2.973 min Scan# 466  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion:101 Resp: 82955  
Ion Ratio Lower Upper  
101 100  
103 65.1 52.1 78.1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#8  
Diethyl Ether  
Concen: 18.51 ug/l  
RT: 3.356 min Scan# 609  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 74 Resp: 35116  
Ion Ratio Lower Upper  
74 100  
45 103.9 46.4 139.1









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 18.26 ug/l  
RT: 3.697 min Scan# 736  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion:101 Resp: 51093  
Ion Ratio Lower Upper  
101 100  
85 42.2 34.6 51.8  
151 79.8 63.5 95.3

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#10  
Methyl Iodide  
Concen: 9.91 ug/l  
RT: 3.887 min Scan# 807  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion:142 Resp: 43574  
Ion Ratio Lower Upper  
142 100  
127 37.3 31.7 47.5  
141 14.5 10.9 16.3





#11





Tert butyl alcohol  
Concen: 101.42 ug/l  
RT: 4.703 min Scan# 1111  
Delta R.T. -0.002 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 59 Resp: 45484  
Ion Ratio Lower Upper  
59 100  
57 10.1 8.0 12.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02

Manual Integrations  
APPROVED




MMDadoda  
4/23/2021 1:15:33 PM




#12

1,1-Dichloroethene  
Concen: 18.18 ug/l  
RT: 3.673 min Scan# 727  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion: 96 Resp: 51340  
Ion Ratio Lower Upper  
96 100  
61 158.7 122.5 183.7  
98 63.9 50.8 76.2





#13  
Acrolein  
Concen: 70.90 ug/l  
RT: 3.549 min Scan# 681  
Delta R.T. 0.002 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 56 Resp: 35134  
Ion Ratio Lower Upper  
56 100  
55 74.1 56.0 84.0

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#14  
Allyl chloride  
Concen: 18.42 ug/l  
RT: 4.247 min Scan# 941  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50



Tgt Ion: 41 Resp: 99693  
Ion Ratio Lower Upper  
41 100  
39 60.5 53.0 79.4  
76 31.2 28.9 43.3






#15  
Acrylonitrile  
Concen: 94.37 ug/l  
RT: 4.896 min Scan# 1183  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02

Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM







Tgt Ion: 53 Resp: 144271  
Ion Ratio Lower Upper  
53 100  
52 82.2 66.1 99.1  
51 35.6 27.9 41.9



#16  
Acetone  
Concen: 85.11 ug/l  
RT: 3.756 min Scan# 758  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion: 43 Resp: 112761  
Ion Ratio Lower Upper  
43 100  
58 33.5 27.0 40.6





#17  
Carbon Disulfide  
Concen: 18.00 ug/l  
RT: 3.984 min Scan# 843  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 76 Resp: 163248  
Ion Ratio Lower Upper  
76 100  
78 9.1 7.8 11.6

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#18  
Methyl Acetate  
Concen: 17.96 ug/l  
RT: 4.252 min Scan# 943  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50



Tgt Ion: 43 Resp: 66001  
Ion Ratio Lower Upper  
43 100  
74 23.5 19.5 29.3





#19





Methyl tert-butyl Ether  
Concen: 18.86 ug/l  
RT: 4.952 min Scan# 1204  
Delta R.T. -0.005 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 73 Resp: 173993  
Ion Ratio Lower Upper  
73 100  
57 23.3 17.3 25.9

Instrument :  
MSVOA\_N  
ClientSampled :  
VN0422WBS02

Manual Integrations  
APPROVED




MMDadoda  
4/23/2021 1:15:33 PM




#20

Methylene Chloride  
Concen: 16.29 ug/l  
RT: 4.469 min Scan# 1024  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion: 84 Resp: 64174  
Ion Ratio Lower Upper  
84 100  
49 123.4 94.6 141.8  
51 39.6 29.9 44.9  
86 63.3 51.4 77.2





#21  
trans-1,2-Dichloroethene  
Concen: 18.43 ug/l  
RT: 4.952 min Scan# 1204  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 96 Resp: 57174  
Ion Ratio Lower Upper  
96 100  
61 140.6 104.9 157.3  
98 64.7 51.8 77.8

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#22  
Diisopropyl ether  
Concen: 18.64 ug/l  
RT: 5.864 min Scan# 1544  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50



Tgt Ion: 45 Resp: 204590  
Ion Ratio Lower Upper  
45 100  
43 65.9 48.9 73.3  
87 25.0 23.7 35.5  
59 10.8 9.9 14.9









#23  
 Vinyl Acetate  
 Concen: 94.07 ug/l  
 RT: 5.805 min Scan# 1522  
 Delta R.T. -0.000 min  
 Lab File: VN066716.D  
 Acq: 22 Apr 2021 15:50

Tgt Ion: 43 Resp: 832359  
 Ion Ratio Lower Upper  
 43 100  
 86 9.7 8.6 13.0

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0422WBS02




Manual Integrations  
 APPROVED


MMDadoda  
 4/23/2021 1:15:33 PM



#24  
 1,1-Dichloroethane  
 Concen: 18.27 ug/l  
 RT: 5.759 min Scan# 1505  
 Delta R.T. 0.000 min  
 Lab File: VN066716.D  
 Acq: 22 Apr 2021 15:50


Tgt Ion: 63 Resp: 111346  
 Ion Ratio Lower Upper  
 63 100  
 98 6.7 3.5 10.4  
 100 4.1 2.3 6.9





#25  
2-Butanone  
Concen: 96.06 ug/l  
RT: 6.746 min Scan# 1873  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampled :  
VN0422WBS02



Tgt Ion: 43 Resp: 182563  
Ion Ratio Lower Upper  
43 100  
72 25.4 23.1 34.7




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#26  
2,2-Dichloropropane  
Concen: 17.70 ug/l  
RT: 6.733 min Scan# 1868  
Delta R.T. -0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion: 77 Resp: 88600  
Ion Ratio Lower Upper  
77 100  
97 23.8 11.9 35.9





#27  
cis-1,2-Dichloroethene  
Concen: 18.21 ug/l  
RT: 6.744 min Scan# 1872  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 96 Resp: 65508  
Ion Ratio Lower Upper  
96 100  
61 148.2 0.0 276.8  
98 64.7 0.0 129.4




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#28  
Bromochloromethane  
Concen: 19.38 ug/l  
RT: 7.111 min Scan# 2009  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 49 Resp: 55560  
Ion Ratio Lower Upper  
49 100  
129 1.6 0.0 3.6  
130 71.7 67.8 101.6









#29  
Tetrahydrofuran  
Concen: 96.56 ug/l  
RT: 7.130 min Scan# 2016  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 42 Resp: 126457  
Ion Ratio Lower Upper  
42 100  
72 42.2 36.6 55.0  
71 39.3 34.3 51.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM




#30  
Chloroform  
Concen: 18.60 ug/l  
RT: 7.293 min Scan# 2077  
Delta R.T. 0.002 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 83 Resp: 106228  
Ion Ratio Lower Upper  
83 100  
85 64.5 52.6 78.8









#31  
Cyclohexane  
Concen: 17.90 ug/l  
RT: 7.578 min Scan# 2183  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50





Tgt Ion: 56 Resp: 103228  
Ion Ratio Lower Upper  
56 100  
69 36.2 26.6 39.8  
84 84.0 70.4 105.6



#32  
1,1,1-Trichloroethane  
Concen: 18.38 ug/l  
RT: 7.494 min Scan# 2152  
Delta R.T. -0.001 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 97 Resp: 87699  
Ion Ratio Lower Upper  
97 100  
99 63.9 50.8 76.2  
61 44.7 34.3 51.5



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02


Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#33  
1,2-Dichloroethane-d4  
Concen: 49.80 ug/l  
RT: 7.950 min Scan# 2322  
Delta R.T. -0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 65 Resp: 176273  
Ion Ratio Lower Upper  
65 100  
67 54.5 0.0 107.2




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#34  
1,4-Difluorobenzene  
Concen: 50.00 ug/l  
RT: 8.516 min Scan# 2533  
Delta R.T. -0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 114 Resp: 470009  
Ion Ratio Lower Upper  
114 100  
63 20.9 0.0 38.6  
88 15.7 0.0 30.6









#35  
Dibromofluoromethane  
Concen: 48.96 ug/l  
RT: 7.513 min Scan# 2159  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 113 Resp: 144107  
Ion Ratio Lower Upper  
113 100  
111 103.6 82.2 123.2  
192 20.9 15.4 23.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#36  
1,1-Dichloropropene  
Concen: 17.80 ug/l  
RT: 7.714 min Scan# 2234  
Delta R.T. -0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 75 Resp: 77788  
Ion Ratio Lower Upper  
75 100  
110 36.3 18.2 54.6  
77 31.5 24.9 37.3









#37  
Ethyl Acetate  
Concen: 17.94 ug/l  
RT: 6.843 min Scan# 1909  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 43    | Resp: | 82384 |
| Ion Ratio | Lower | Upper |       |
| 43        | 100   |       |       |
| 61        | 15.3  | 11.6  | 17.4  |
| 70        | 9.8   | 8.6   | 13.0  |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#38  
Carbon Tetrachloride  
Concen: 17.99 ug/l  
RT: 7.701 min Scan# 2229  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 117   | Resp: | 74816 |
| Ion Ratio | Lower | Upper |       |
| 117       | 100   |       |       |
| 119       | 93.8  | 76.0  | 114.0 |
| 121       | 30.8  | 24.5  | 36.7  |





#39

Methylcyclohexane

Concen: 18.04 ug/l

RT: 9.015 min Scan# 2719

Delta R.T. 0.000 min

Lab File: VN066716.D





Acq: 22 Apr 2021 15:50

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 93458 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 55       | 79.3  | 58.9  | 88.3  |
| 98       | 44.1  | 35.9  | 53.9  |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02

Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#40

Benzene

Concen: 18.19 ug/l




RT: 7.967 min Scan# 2328


Delta R.T. -0.002 min

Lab File: VN066716.D


Acq: 22 Apr 2021 15:50

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 78    | Resp: | 255856 |
| Ion      | Ratio | Lower | Upper  |
| 78       | 100   |       |        |
| 77       | 23.7  | 19.0  | 28.4   |









#41  
Methacrylonitrile  
Concen: 17.57 ug/l  
RT: 7.090 min Scan# 2001  
Delta R.T. 0.006 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50





|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 41    | Resp: | 35852 |
| Ion Ratio | Lower | Upper |       |
| 41        | 100   |       |       |
| 39        | 49.3  | 0.0   | 0.0#  |
| 67        | 85.9  | 0.0   | 0.0#  |
| 52        | 35.9  | 0.0   | 0.0#  |



#42  
1,2-Dichloroethane  
Concen: 17.78 ug/l  
RT: 8.052 min Scan# 2360  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 62    | Resp: | 79833 |
| Ion Ratio | Lower | Upper |       |
| 62        | 100   |       |       |
| 98        | 9.2   | 0.0   | 18.6  |



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02


Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#43  
Isopropyl Acetate  
Concen: 18.57 ug/l  
RT: 8.093 min Scan# 2375  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 43 Resp: 136119  
Ion Ratio Lower Upper  
43 100  
61 20.9 22.5 33.7#  
87 12.6 11.1 16.7




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#44  
Trichloroethene  
Concen: 18.71 ug/l  
RT: 8.768 min Scan# 2627  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 130 Resp: 62607  
Ion Ratio Lower Upper  
130 100  
95 100.1 0.0 199.6









#45  
1,2-Dichloropropane  
Concen: 18.01 ug/l  
RT: 9.053 min Scan# 2733  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 63 Resp: 68876  
Ion Ratio Lower Upper  
63 100  
65 31.8 24.9 37.3

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#46  
Dibromomethane  
Concen: 18.45 ug/l  
RT: 9.141 min Scan# 2766  
Delta R.T. -0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion: 93 Resp: 41187  
Ion Ratio Lower Upper  
93 100  
95 82.9 67.0 100.6  
174 102.4 83.2 124.8





#47  
Bromodichloromethane  
Concen: 17.97 ug/l  
RT: 9.337 min Scan# 2839  
Delta R.T. -0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 83 Resp: 84316  
Ion Ratio Lower Upper  
83 100  
85 63.6 51.0 76.4  
127 8.5 7.0 10.6




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM




#48  
Methyl methacrylate  
Concen: 18.47 ug/l  
RT: 9.136 min Scan# 2764  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 41 Resp: 59379  
Ion Ratio Lower Upper  
41 100  
69 87.7 75.9 113.9  
39 51.1 46.6 69.8









#49  
1,4-Dioxane  
Concen: 445.73 ug/l  
RT: 9.139 min Scan# 2765  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50





Tgt Ion: 88 Resp: 17320  
Ion Ratio Lower Upper  
88 100  
43 35.7 24.7 37.1  
58 72.9 52.6 78.8



#50  
Toluene-d8  
Concen: 49.30 ug/l  
RT: 10.029 min Scan# 3097  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 98 Resp: 565865  
Ion Ratio Lower Upper  
98 100  
100 64.0 50.5 75.7



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02


Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#51  
4-Methyl-2-Pentanone  
Concen: 94.60 ug/l  
RT: 9.922 min Scan# 3057  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 43 Resp: 387893  
Ion Ratio Lower Upper  
43 100  
58 38.7 32.5 48.7




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#52  
Toluene  
Concen: 18.56 ug/l  
RT: 10.093 min Scan# 3121  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion: 92 Resp: 153313  
Ion Ratio Lower Upper  
92 100  
91 174.6 140.7 211.1





#53  
t-1,3-Dichloropropene  
Concen: 18.28 ug/l  
RT: 10.321 min Scan# 3206  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 75 Resp: 90535  
Ion Ratio Lower Upper  
75 100  
77 31.9 26.2 39.4




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM




#54  
cis-1,3-Dichloropropene  
Concen: 18.24 ug/l  
RT: 9.777 min Scan# 3003  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion: 75 Resp: 102944  
Ion Ratio Lower Upper  
75 100  
77 31.5 24.9 37.3  
39 47.9 36.2 54.4









#55  
1,1,2-Trichloroethane  
Concen: 18.03 ug/l  
RT: 10.501 min Scan# 3273  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50





|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 97    | Resp: | 58947 |
| Ion      | Ratio | Lower | Upper |
| 97       | 100   |       |       |
| 83       | 90.7  | 67.7  | 101.5 |
| 85       | 57.9  | 44.5  | 66.7  |
| 99       | 65.3  | 49.4  | 74.0  |



#56  
Ethyl methacrylate  
Concen: 17.62 ug/l  
RT: 10.372 min Scan# 3225  
Delta R.T. 0.002 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 69    | Resp: | 80109 |
| Ion      | Ratio | Lower | Upper |
| 69       | 100   |       |       |
| 41       | 64.9  | 49.6  | 74.4  |
| 39       | 33.2  | 28.0  | 42.0  |



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02


Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#57  
1,3-Dichloropropane  
Concen: 18.44 ug/l  
RT: 10.646 min Scan# 3327  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 76 Resp: 101474  
Ion Ratio Lower Upper  
76 100  
78 31.9 25.7 38.5

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#58  
2-Chloroethyl Vinyl ether  
Concen: 167.32 ug/l  
RT: 9.632 min Scan# 2949  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 63 Resp: 139669  
Ion Ratio Lower Upper  
63 100  
106 25.9 22.6 33.8





#59  
2-Hexanone  
Concen: 87.56 ug/l  
RT: 10.694 min Scan# 3345  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 43 Resp: 233798  
Ion Ratio Lower Upper  
43 100  
58 55.1 28.1 84.3

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#60  
Dibromochloromethane  
Concen: 18.37 ug/l  
RT: 10.842 min Scan# 3400  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 129 Resp: 63847  
Ion Ratio Lower Upper  
129 100  
127 77.2 38.5 115.5





#61  
1,2-Dibromoethane  
Concen: 18.41 ug/l  
RT: 10.946 min Scan# 3439  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:107 Resp: 59161  
Ion Ratio Lower Upper  
107 100  
109 93.9 75.6 113.4

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#62  
4-Bromofluorobenzene  
Concen: 47.99 ug/l  
RT: 12.346 min Scan# 3961  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 95 Resp: 188657  
Ion Ratio Lower Upper  
95 100  
174 84.7 0.0 169.0  
176 82.0 0.0 162.4





#63  
Chlorobenzene-d5  
Concen: 50.00 ug/l  
RT: 11.348 min Scan# 3589  
Delta R.T. -0.001 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:117 Resp: 429268  
Ion Ratio Lower Upper  
117 100  
82 54.5 43.7 65.5  
119 31.7 26.1 39.1

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#64  
Tetrachloroethene  
Concen: 18.85 ug/l  
RT: 10.571 min Scan# 3299  
Delta R.T. -0.002 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion:164 Resp: 57129  
Ion Ratio Lower Upper  
164 100  
166 127.3 102.6 154.0  
129 94.8 76.2 114.4  
131 89.7 73.4 110.2





#65  
Chlorobenzene  
Concen: 18.25 ug/l  
RT: 11.375 min Scan# 3599  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:112 Resp: 159333  
Ion Ratio Lower Upper  
112 100  
114 31.5 26.1 39.1

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#66  
1,1,1,2-Tetrachloroethane  
Concen: 18.59 ug/l  
RT: 11.450 min Scan# 3627  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion:131 Resp: 58911  
Ion Ratio Lower Upper  
131 100  
133 96.4 48.0 144.0  
119 70.0 34.1 102.2





#67  
Ethyl Benzene  
Concen: 18.49 ug/l  
RT: 11.453 min Scan# 3628  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion: 91 Resp: 277417  
Ion Ratio Lower Upper  
91 100  
106 31.1 25.5 38.3




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#68  
m/p-Xylenes  
Concen: 37.17 ug/l  
RT: 11.563 min Scan# 3669  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion: 106 Resp: 207881  
Ion Ratio Lower Upper  
106 100  
91 203.5 161.4 242.0





#69  
o-Xylene  
Concen: 18.74 ug/l  
RT: 11.890 min Scan# 3791  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02



Tgt Ion:106 Resp: 103247  
Ion Ratio Lower Upper  
106 100  
91 210.6 106.8 320.4




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#70  
Styrene  
Concen: 17.34 ug/l  
RT: 11.909 min Scan# 3798  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50


Tgt Ion:104 Resp: 158193  
Ion Ratio Lower Upper  
104 100  
78 49.0 39.0 58.4  
103 55.1 44.5 66.7





#71  
Bromoform  
Concen: 18.31 ug/l  
RT: 12.073 min Scan# 3859  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:173 Resp: 44003  
Ion Ratio Lower Upper  
173 100  
175 47.4 23.8 71.2  
254 0.0 0.0 0.0

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.00 ug/l  
RT: 13.288 min Scan# 4312  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion:152 Resp: 192843  
Ion Ratio Lower Upper  
152 100  
115 57.0 28.5 85.6  
150 164.2 0.0 347.6





#73  
Isopropylbenzene  
Concen: 19.46 ug/l  
RT: 12.193 min Scan# 3904  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:105 Resp: 266757  
Ion Ratio Lower Upper  
105 100  
120 26.4 13.3 39.9

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#74  
N-ethyl acetate  
Concen: 9.08 ug/l  
RT: 12.035 min Scan# 3845  
Delta R.T. 0.016 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 43 Resp: 29616  
Ion Ratio Lower Upper  
43 100  
70 29.2 34.4 51.6#  
55 19.1 20.7 31.1#  
61 19.0 19.1 28.7#





#75  
1,1,2,2-Tetrachloroethane  
Concen: 18.39 ug/l  
RT: 12.448 min Scan# 3999  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 83 Resp: 85048  
Ion Ratio Lower Upper  
83 100  
131 9.6 5.3 15.8  
85 66.8 32.0 96.0

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#76  
1,2,3-Trichloropropane  
Concen: 19.08 ug/l m  
RT: 12.499 min Scan# 4018  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 75 Resp: 78746  
Ion Ratio Lower Upper  
75 100  
77 0.0 0.0 0.0





#77  
Bromobenzene  
Concen: 19.11 ug/l  
RT: 12.470 min Scan# 4007  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:156 Resp: 67760  
Ion Ratio Lower Upper  
156 100  
77 193.4 96.3 288.8  
158 97.5 49.0 147.0

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#78  
n-propylbenzene  
Concen: 19.50 ug/l  
RT: 12.534 min Scan# 4031  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 91 Resp: 297931  
Ion Ratio Lower Upper  
91 100  
120 23.3 11.8 35.3





#79  
2-Chlorotoluene  
Concen: 18.60 ug/l  
RT: 12.617 min Scan# 4062  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 91 Resp: 184262  
Ion Ratio Lower Upper  
91 100  
126 34.7 17.4 52.3

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#80  
1,3,5-Trimethylbenzene  
Concen: 19.16 ug/l  
RT: 12.679 min Scan# 4085  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 105 Resp: 223181  
Ion Ratio Lower Upper  
105 100  
120 48.7 24.3 72.8









#81  
trans-1,4-Dichloro-2-butene  
Concen: 15.65 ug/l  
RT: 12.250 min Scan# 3925  
Delta R.T. 0.006 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50





Tgt Ion: 75 Resp: 15384  
Ion Ratio Lower Upper  
75 100  
53 86.9 74.6 111.8  
89 51.1 35.8 53.8



#82  
4-Chlorotoluene  
Concen: 19.08 ug/l  
RT: 12.719 min Scan# 4100  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 91 Resp: 179276  
Ion Ratio Lower Upper  
91 100  
126 33.2 17.2 51.6



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02


Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#83  
tert-Butylbenzene  
Concen: 18.03 ug/l  
RT: 12.936 min Scan# 4181  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampled :  
VN0422WBS02




Tgt Ion:119 Resp: 187055  
Ion Ratio Lower Upper  
119 100  
91 66.4 33.0 98.9  
134 27.1 12.7 38.1

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#84  
1,2,4-Trimethylbenzene  
Concen: 19.02 ug/l  
RT: 12.984 min Scan# 4199  
Delta R.T. -0.001 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion:105 Resp: 211949  
Ion Ratio Lower Upper  
105 100  
120 46.9 22.4 67.0





#85  
 sec-Butylbenzene  
 Concen: 18.99 ug/l  
 RT: 13.116 min Scan# 4248  
 Delta R.T. -0.000 min  
 Lab File: VN066716.D  
 Acq: 22 Apr 2021 15:50




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0422WBS02




Tgt Ion:105 Resp: 247624  
 Ion Ratio Lower Upper  
 105 100  
 134 20.2 10.0 30.0

Manual Integrations  
 APPROVED



MMDadoda  
 4/23/2021 1:15:33 PM




#86  
 p-Isopropyltoluene  
 Concen: 19.90 ug/l  
 RT: 13.231 min Scan# 4291  
 Delta R.T. -0.003 min  
 Lab File: VN066716.D  
 Acq: 22 Apr 2021 15:50




Tgt Ion:119 Resp: 215888  
 Ion Ratio Lower Upper  
 119 100  
 134 26.8 13.3 39.8  
 91 23.1 11.8 35.3





#87  
1,3-Dichlorobenzene  
Concen: 18.83 ug/l  
RT: 13.229 min Scan# 4290  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:146 Resp: 112395  
Ion Ratio Lower Upper  
146 100  
111 39.0 19.5 58.5  
148 63.8 32.3 96.9

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#88  
1,4-Dichlorobenzene  
Concen: 17.25 ug/l  
RT: 13.306 min Scan# 4319  
Delta R.T. 0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion:146 Resp: 110698  
Ion Ratio Lower Upper  
146 100  
111 38.7 19.0 57.0  
148 66.8 31.8 95.3





#89  
n-Butylbenzene  
Concen: 18.00 ug/l  
RT: 13.561 min Scan# 4414  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion: 91 Resp: 173894  
Ion Ratio Lower Upper  
91 100  
92 51.8 26.1 78.3  
134 24.6 13.0 38.9

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#90  
Hexachloroethane  
Concen: 15.75 ug/l  
RT: 13.816 min Scan# 4509  
Delta R.T. -0.000 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Tgt Ion: 117 Resp: 38208  
Ion Ratio Lower Upper  
117 100  
201 83.7 37.8 113.3





#91  
1,2-Dichlorobenzene  
Concen: 18.52 ug/l  
RT: 13.599 min Scan# 4428  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Tgt Ion:146 Resp: 112896  
Ion Ratio Lower Upper  
146 100  
111 39.8 20.1 60.3  
148 63.3 32.0 96.0

Manual Integrations  
APPROVED



MMDadoda  
4/23/2021 1:15:33 PM




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 17.13 ug/l  
RT: 14.221 min Scan# 4660  
Delta R.T. 0.005 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50



Tgt Ion: 75 Resp: 14128  
Ion Ratio Lower Upper  
75 100  
155 80.1 44.1 132.4  
157 99.0 56.0 168.2









#93  
1,2,4-Trichlorobenzene  
Concen: 19.41 ug/l  
RT: 14.854 min Scan# 4896  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion:180 Resp: 66083  
Ion Ratio Lower Upper  
180 100  
182 92.1 47.3 142.0  
145 27.8 15.6 46.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02




Manual Integrations  
APPROVED


MMDadoda  
4/23/2021 1:15:33 PM



#94  
Hexachlorobutadiene  
Concen: 19.31 ug/l  
RT: 14.950 min Scan# 4932  
Delta R.T. 0.002 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

Tgt Ion:225 Resp: 36341  
Ion Ratio Lower Upper  
225 100  
223 63.4 31.9 95.9  
227 63.1 31.6 95.0









#95  
Naphthalene  
Concen: 20.17 ug/l  
RT: 15.074 min Scan# 4978  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 128   | Resp: | 169041 |
| Ion Ratio | Lower | Upper |        |
| 128       | 100   |       |        |
| 127       | 12.6  | 10.2  | 15.2   |
| 129       | 10.6  | 8.6   | 13.0   |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0422WBS02

Manual Integrations  
APPROVED

MMDadoda  
4/23/2021 1:15:33 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 19.26 ug/l  
RT: 15.251 min Scan# 5044  
Delta R.T. 0.003 min  
Lab File: VN066716.D  
Acq: 22 Apr 2021 15:50

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 180   | Resp: | 62908 |
| Ion Ratio | Lower | Upper |       |
| 180       | 100   |       |       |
| 182       | 98.5  | 47.4  | 142.3 |
| 145       | 32.1  | 16.5  | 49.5  |

