


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042221\
 Data File : VN066705.D
 Acq On : 22 Apr 2021 9:09
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042021W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Wed Apr 21 16:55:39 2021



AutoFind: Averaged scan 3948 to 3990; Bkg corrected with scan 3947

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	1699	PASS
75	95	30	60	49.6	4345	PASS
95	95	100	100	100.0	8762	PASS
96	95	5	9	7.7	673	PASS
173	174	0.00	2	0.6	44	PASS
174	95	50	100	85.3	7472	PASS
175	174	5	9	7.4	556	PASS
176	174	95	101	96.2	7191	PASS
177	176	5	9	6.7	483	PASS