

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
 Data File : VN072147.D
 Acq On : 22 Apr 2022 11:01
 Operator : JC\MD
 Sample : VN0422WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0422WBL01

Quant Time: Apr 23 03:12:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	438923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	741923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	768836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	271048	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	238686	47.634	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.260%
35) Dibromofluoromethane	8.016	113	219351	50.243	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.480%
50) Toluene-d8	10.439	98	918592	51.435	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.860%
62) 4-Bromofluorobenzene	12.727	95	301145	52.137	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
Data File : VN072147.D
Acq On : 22 Apr 2022 11:01
Operator : JC\MD
Sample : VN0422WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0422WBL01

Quant Time: Apr 23 03:12:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

