

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
 Data File : VN072153.D
 Acq On : 22 Apr 2022 13:46
 Operator : JC\MD
 Sample : N2482-19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2022-04-19

Quant Time: Apr 23 04:43:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	423275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	662553	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	562089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	183380	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	245836	50.875	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.740%	
35) Dibromofluoromethane	8.022	113	218560	56.059	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 112.120%	
50) Toluene-d8	10.439	98	777868	48.773	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.540%	
62) 4-Bromofluorobenzene	12.727	95	215638	41.806	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 83.620%	
Target Compounds						Qvalue
16) Acetone	4.292	43	41749	20.216	ug/l	97
25) 2-Butanone	7.333	43	10649	3.577	ug/l	97
29) Tetrahydrofuran	7.692	42	4589	2.381	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
Data File : VN072153.D
Acq On : 22 Apr 2022 13:46
Operator : JC\MD
Sample : N2482-19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-2022-04-19

Quant Time: Apr 23 04:43:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

