

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
 Data File : VN072156.D
 Acq On : 22 Apr 2022 14:58
 Operator : JC\MD
 Sample : N2526-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-5S

Quant Time: Apr 23 03:18:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	567198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	952037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1075156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	390990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	296651	45.813	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.620%
35) Dibromofluoromethane	8.022	113	273753	48.865	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.740%
50) Toluene-d8	10.439	98	1240738	54.140	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.280%
62) 4-Bromofluorobenzene	12.727	95	436814	58.935	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.860%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	23847	8.617	ug/l #	88
20) Methylene Chloride	5.098	84	41546	7.123	ug/l	93
43) Isopropyl Acetate	8.557	43	42810	3.006	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
Data File : VN072156.D
Acq On : 22 Apr 2022 14:58
Operator : JC\MD
Sample : N2526-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-5S

Quant Time: Apr 23 03:18:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

