

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
 Data File : VN072163.D
 Acq On : 22 Apr 2022 17:47
 Operator : JC\MD
 Sample : N2494-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-1D

Quant Time: Apr 23 03:22:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	531322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	910422	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	999575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	368412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	273982	45.169	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.340%
35) Dibromofluoromethane	8.022	113	256554	47.889	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.780%
50) Toluene-d8	10.439	98	1174851	53.608	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.220%
62) 4-Bromofluorobenzene	12.727	95	407252	57.458	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	114.920%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	22429	8.652	ug/l	95
20) Methylene Chloride	5.093	84	73379	13.430	ug/l	90
43) Isopropyl Acetate	8.557	43	50807	3.730	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
Data File : VN072163.D
Acq On : 22 Apr 2022 17:47
Operator : JC\MD
Sample : N2494-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-1D

Quant Time: Apr 23 03:22:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

