

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
 Data File : VN072166.D
 Acq On : 22 Apr 2022 18:59
 Operator : JC\MD
 Sample : N2522-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NWB-1757

Quant Time: Apr 23 03:23:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	597800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1029784	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1106900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	473548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	314115	46.027	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.060%
35) Dibromofluoromethane	8.022	113	292273	48.232	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.460%
50) Toluene-d8	10.439	98	1287728	51.948	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.900%
62) 4-Bromofluorobenzene	12.727	95	490905	61.232	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	122.460%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	142880	48.989	ug/l	92
25) 2-Butanone	7.339	43	9021	2.146	ug/l	93
52) Toluene	10.504	92	5197	0.308	ug/l	98
68) m/p-Xylenes	11.951	106	7960	0.552	ug/l	96
69) o-Xylene	12.280	106	3885	0.272	ug/l	88
84) 1,2,4-Trimethylbenzene	13.363	105	16985	0.551	ug/l	94
95) Naphthalene	15.504	128	28168	5.010	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
Data File : VN072166.D
Acq On : 22 Apr 2022 18:59
Operator : JC\MD
Sample : N2522-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NWB-1757

Quant Time: Apr 23 03:23:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

