

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
 Data File : VN072168.D
 Acq On : 22 Apr 2022 19:48
 Operator : JC\MD
 Sample : PB144286TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144286TB

Quant Time: Apr 23 03:23:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	626071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1073461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1149479	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	451768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	324554	45.409	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.820%
35) Dibromofluoromethane	8.022	113	299737	47.451	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.900%
50) Toluene-d8	10.439	98	1380803	53.436	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.880%
62) 4-Bromofluorobenzene	12.727	95	491417	58.802	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.600%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	18712	6.126	ug/l #	87
20) Methylene Chloride	5.098	84	37119	5.765	ug/l	93
43) Isopropyl Acetate	8.557	43	65686	4.090	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
Data File : VN072168.D
Acq On : 22 Apr 2022 19:48
Operator : JC\MD
Sample : PB144286TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144286TB

Quant Time: Apr 23 03:23:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

