

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
 Data File : VN072171.D
 Acq On : 22 Apr 2022 21:00
 Operator : JC\MD
 Sample : N2524-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 33656-57

Quant Time: Apr 23 03:24:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	624366	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1062835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1135399	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	443708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	322280	45.214	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.420%
35) Dibromofluoromethane	8.022	113	300880	48.109	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.220%
50) Toluene-d8	10.439	98	1328372	51.921	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.840%
62) 4-Bromofluorobenzene	12.727	95	472603	57.116	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	114.240%
Target Compounds						
					Qvalue	
40) Benzene	8.457	78	76171	2.761	ug/l	99
52) Toluene	10.498	92	50719	2.908	ug/l	97
95) Naphthalene	15.504	128	23688	4.886	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042222\
Data File : VN072171.D
Acq On : 22 Apr 2022 21:00
Operator : JC\MD
Sample : N2524-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
33656-57

Quant Time: Apr 23 03:24:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

