

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN042319\
 Data File : VN055193.D
 Acq On : 23 Apr 2019 9:42
 Operator : JC/SP
 Sample : VN0423MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0423MBL01

Quant Time: Apr 23 15:37:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	392443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	649740	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	580211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	156415	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	257196	48.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.84%	
35) Dibromofluoromethane	7.59	113	211337	47.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.76%	
50) Toluene-d8	10.09	98	796011	48.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.34%	
62) 4-Bromofluorobenzene	12.40	95	238114	45.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.96%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042319\
Data File : VN055193.D
Acq On : 23 Apr 2019 9:42
Operator : JC/SP
Sample : VN0423MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0423MBL01

Quant Time: Apr 23 15:37:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

