

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042319\
 Data File : VN055198.D
 Acq On : 23 Apr 2019 11:49
 Operator : JC/SP
 Sample : K2532-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Apr 24 03:19:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	370707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	625561	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	571096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	166400	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	251548	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
35) Dibromofluoromethane	7.59	113	207421	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	10.09	98	770236	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	12.40	95	237824	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
Target Compounds						
16) Acetone	3.83	43	10552	6.361	ug/l	98
20) Methylene Chloride	4.55	84	43284	9.077	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042319\
Data File : VN055198.D
Acq On : 23 Apr 2019 11:49
Operator : JC/SP
Sample : K2532-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Apr 24 03:19:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

