

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042319\
 Data File : VN055199.D
 Acq On : 23 Apr 2019 12:14
 Operator : JC/SP
 Sample : K2532-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-14

Quant Time: Apr 24 03:57:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	370379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	640099	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	572748	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	172332	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	254684	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
35) Dibromofluoromethane	7.59	113	206274	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
50) Toluene-d8	10.09	98	773880	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
62) 4-Bromofluorobenzene	12.40	95	237619	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
Target Compounds						
16) Acetone	3.82	43	9903	5.976	ug/l	93
20) Methylene Chloride	4.55	84	697962	146.496	ug/l	99
43) Isopropyl Acetate	8.17	43	14636	1.885	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042319\
Data File : VN055199.D
Acq On : 23 Apr 2019 12:14
Operator : JC/SP
Sample : K2532-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-14

Quant Time: Apr 24 03:57:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

