

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042319\
 Data File : VN055200.D
 Acq On : 23 Apr 2019 12:39
 Operator : JC/SP
 Sample : K2532-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-17

Quant Time: Apr 24 04:12:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	376835	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	645957	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	592878	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	187793	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	258153	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
35) Dibromofluoromethane	7.59	113	214023	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	10.09	98	796181	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
62) 4-Bromofluorobenzene	12.40	95	251127	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
Target Compounds						
16) Acetone	3.82	43	9040m	5.361	ug/l	
20) Methylene Chloride	4.55	84	108302	22.342	ug/l	97
30) Chloroform	7.37	83	14094	1.803	ug/l	94
43) Isopropyl Acetate	8.17	43	13682	1.746	ug/l	95
95) Naphthalene	15.13	128	29057	7.877	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042319\
Data File : VN055200.D
Acq On : 23 Apr 2019 12:39
Operator : JC/SP
Sample : K2532-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-17

Quant Time: Apr 24 04:12:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

