

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042319\
 Data File : VN055213.D
 Acq On : 23 Apr 2019 18:11
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 24 04:42:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.644	0.674	-4.7	96	0.00
3 P	Chloromethane	0.789	0.776	1.6	94	0.00
4 C	Vinyl Chloride	0.632	0.611	3.3#	93	0.00
5 T	Bromomethane	0.367	0.318	13.4	92	0.00
6 T	Chloroethane	0.333	0.316	5.1	94	0.00
7 T	Trichlorofluoromethane	0.860	0.836	2.8	95	0.00
8 T	Diethyl Ether	0.307	0.334	-8.8	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.537	0.512	4.7	96	0.00
10 T	Methyl Iodide	0.753	0.793	-5.3	97	0.00
11 T	Tert butyl alcohol	0.040	0.052	-30.0#	130	0.00
12 CM	1,1-Dichloroethene	0.514	0.509	1.0#	97	0.00
13 T	Acrolein	0.054	0.062	-14.8	109	0.00
14 T	Allyl chloride	1.017	1.031	-1.4	99	0.00
15 T	Acrylonitrile	0.217	0.252	-16.1	112	0.00
16 T	Acetone	0.224	0.205	8.5	81	0.00
17 T	Carbon Disulfide	1.438	1.355	5.8	89	0.00
18 T	Methyl Acetate	0.605	0.582	3.8	121	0.00
19 T	Methyl tert-butyl Ether	1.422	1.562	-9.8	107	0.00
20 T	Methylene Chloride	0.643	0.606	5.8	99	0.00
21 T	trans-1,2-Dichloroethene	0.539	0.536	0.6	96	0.00
22 T	Diisopropyl ether	2.027	2.087	-3.0	101	0.00
23 T	Vinyl Acetate	1.362	1.596	-17.2	107	0.00
24 P	1,1-Dichloroethane	1.090	1.078	1.1	98	0.00
25 T	2-Butanone	0.289	0.322	-11.4	105	0.00
26 T	2,2-Dichloropropane	0.738	0.719	2.6	97	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.616	1.0	98	0.00
28 T	Bromochloromethane	0.556	0.550	1.1	99	0.00
29 T	Tetrahydrofuran	0.178	0.213	-19.7	116	0.00
30 C	Chloroform	1.037	1.011	2.5#	97	0.00
31 T	Cyclohexane	1.252	1.009	19.4	96	0.00
32 T	1,1,1-Trichloroethane	0.833	0.846	-1.6	99	0.00
33 S	1,2-Dichloroethane-d4	0.677	0.676	0.1	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.340	0.336	1.2	97	0.00
36 T	1,1-Dichloropropene	0.464	0.484	-4.3	96	0.00
37 T	Ethyl Acetate	0.375	0.443	-18.1	114	0.00
38 T	Carbon Tetrachloride	0.435	0.449	-3.2	97	0.00
39 T	Methylcyclohexane	0.554	0.534	3.6	95	0.00
40 TM	Benzene	1.433	1.471	-2.7	99	0.00
41 T	Methacrylonitrile	0.196	0.191	2.6	87	0.00
42 TM	1,2-Dichloroethane	0.500	0.512	-2.4	100	0.00
43 T	Isopropyl Acetate	0.606	0.755	-24.6	116	0.00
44 TM	Trichloroethene	0.362	0.363	-0.3	97	0.00
45 C	1,2-Dichloropropane	0.404	0.407	-0.7#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042319\
 Data File : VN055213.D
 Acq On : 23 Apr 2019 18:11
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 24 04:42:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.231	0.236	-2.2	100	0.00
47 T	Bromodichloromethane	0.463	0.481	-3.9	97	0.00
48 T	Methyl methacrylate	0.294	0.350	-19.0	114	0.00
49 T	1,4-Dioxane	0.003	0.004	-33.3#	113	0.00
50 S	Toluene-d8	1.259	1.242	1.4	98	0.00
51 T	4-Methyl-2-Pentanone	0.349	0.423	-21.2	116	0.00
52 CM	Toluene	0.840	0.862	-2.6#	97	0.00
53 T	t-1,3-Dichloropropene	0.425	0.486	-14.4	103	0.00
54 T	cis-1,3-Dichloropropene	0.528	0.579	-9.7	100	0.00
55 T	1,1,2-Trichloroethane	0.310	0.327	-5.5	103	0.00
56 T	Ethyl methacrylate	0.386	0.477	-23.6	111	0.00
57 T	1,3-Dichloropropane	0.539	0.580	-7.6	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.124	0.148	-19.4	100	0.00
59 T	2-Hexanone	0.220	0.269	-22.3	110	0.00
60 T	Dibromochloromethane	0.319	0.346	-8.5	101	0.00
61 T	1,2-Dibromoethane	0.291	0.325	-11.7	105	0.00
62 S	4-Bromofluorobenzene	0.399	0.411	-3.0	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.408	0.400	2.0	100	0.00
65 PM	Chlorobenzene	1.021	0.986	3.4	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.378	0.380	-0.5	99	0.00
67 C	Ethyl Benzene	1.818	1.823	-0.3#	97	0.00
68 T	m/p-Xylenes	0.652	0.665	-2.0	97	0.00
69 T	o-Xylene	0.647	0.656	-1.4	98	0.00
70 T	Styrene	1.003	1.086	-8.3	97	0.00
71 P	Bromoform	0.226	0.253	-11.9	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	4.800	4.154	13.5	99	0.00
74 T	N-amyl acetate	1.289	1.532	-18.9	114	0.00
75 P	1,1,2,2-Tetrachloroethane	1.270	1.149	9.5	107	0.00
76 T	1,2,3-Trichloropropane	1.014	0.857	15.5	92	0.00
77 T	Bromobenzene	1.104	1.021	7.5	99	0.00
78 T	n-propylbenzene	5.014	4.605	8.2	97	0.00
79 T	2-Chlorotoluene	3.166	2.848	10.0	98	0.00
80 T	1,3,5-Trimethylbenzene	3.743	3.447	7.9	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.268	0.285	-6.3	112	0.00
82 T	4-Chlorotoluene	2.913	2.728	6.4	96	0.00
83 T	tert-Butylbenzene	3.348	2.899	13.4	98	0.00
84 T	1,2,4-Trimethylbenzene	3.532	3.383	4.2	98	0.00
85 T	sec-Butylbenzene	4.235	3.769	11.0	97	0.00
86 T	p-Isopropyltoluene	3.489	3.295	5.6	97	0.00
87 T	1,3-Dichlorobenzene	1.694	1.662	1.9	97	0.00
88 T	1,4-Dichlorobenzene	1.650	1.594	3.4	97	0.00
89 T	n-Butylbenzene	2.586	2.665	-3.1	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042319\
Data File : VN055213.D
Acq On : 23 Apr 2019 18:11
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 24 04:42:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.619	0.540	12.8	96	0.00
91 T	1,2-Dichlorobenzene	1.740	1.625	6.6	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.152	0.165	-8.6	112	0.00
93 T	1,2,4-Trichlorobenzene	0.614	0.713	-16.1	99	0.00
94 T	Hexachlorobutadiene	0.623	0.548	12.0	100	0.00
95 T	Naphthalene	1.298	1.606	-23.7	105	0.00
96 T	1,2,3-Trichlorobenzene	0.639	0.717	-12.2	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6