

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
 Data File : VN066734.D
 Acq On : 23 Apr 2021 13:43
 Operator : JC/MD
 Sample : VN0423MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0423MBL01

Quant Time: Apr 26 03:16:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	302287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	461746	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	404186	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	173880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	175246	47.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.90%
35) Dibromofluoromethane	7.513	113	145956	50.48	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.96%
50) Toluene-d8	10.029	98	559928	49.65	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.30%
62) 4-Bromofluorobenzene	12.346	95	167614	43.40	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
Data File : VN066734.D
Acq On : 23 Apr 2021 13:43
Operator : JC/MD
Sample : VN0423MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0423MBL01

Quant Time: Apr 26 03:16:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

