

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
 Data File : VN066742.D
 Acq On : 23 Apr 2021 17:00
 Operator : JC/MD
 Sample : PB135928TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135928TB

Quant Time: Apr 26 09:11:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	315699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	481568	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	441117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.288	152	188394	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	179637	46.58	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.16%		
35) Dibromofluoromethane	7.513	113	149490	49.57	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.14%		
50) Toluene-d8	10.029	98	594839	50.58	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.16%		
62) 4-Bromofluorobenzene	12.346	95	192458	47.78	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	95.56%		
Target Compounds					Qvalue	
16) Acetone	3.756	43	63005	43.64	ug/l	95
20) Methylene Chloride	4.469	84	23773	5.54	ug/l	98
43) Isopropyl Acetate	8.093	43	28453	3.79	ug/l #	89
95) Naphthalene	15.084	128	21755	4.26	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
Data File : VN066742.D
Acq On : 23 Apr 2021 17:00
Operator : JC/MD
Sample : PB135928TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135928TB

Quant Time: Apr 26 09:11:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

