

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
 Data File : VN066745.D
 Acq On : 23 Apr 2021 18:10
 Operator : JC/MD
 Sample : PB135928ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135928ZHE#03

Quant Time: Apr 26 09:12:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	330180	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	508077	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	473402	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	198816	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	191339	47.43	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.86%
35) Dibromofluoromethane	7.516	113	159670	50.18	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.36%
50) Toluene-d8	10.029	98	630569	50.82	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.64%
62) 4-Bromofluorobenzene	12.346	95	202531	47.66	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.32%
Target Compounds						
					Qvalue	
16) Acetone	3.759	43	33034	21.88	ug/l	100
20) Methylene Chloride	4.472	84	16791	3.74	ug/l	98
43) Isopropyl Acetate	8.095	43	30175	3.81	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
Data File : VN066745.D
Acq On : 23 Apr 2021 18:10
Operator : JC/MD
Sample : PB135928ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135928ZHE#03

Quant Time: Apr 26 09:12:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

