

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
 Data File : VN066748.D
 Acq On : 23 Apr 2021 19:21
 Operator : JC/MD
 Sample : M2132-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2-COMP

Quant Time: Apr 26 09:13:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	311874	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	483279	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	447513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	194493	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	179420	47.09	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.18%
35) Dibromofluoromethane	7.513	113	152209	50.29	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.58%
50) Toluene-d8	10.029	98	600526	50.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.76%
62) 4-Bromofluorobenzene	12.346	95	191458	47.37	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.74%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	52968	37.14	ug/l	98
18) Methyl Acetate	4.260	43	5129	1.30	ug/l #	55
20) Methylene Chloride	4.475	84	23850	5.62	ug/l	95
43) Isopropyl Acetate	8.095	43	27494	3.65	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042321\
Data File : VN066748.D
Acq On : 23 Apr 2021 19:21
Operator : JC/MD
Sample : M2132-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2-COMP

Quant Time: Apr 26 09:13:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

