

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042325\
 Data File : VN086387.D
 Acq On : 23 Apr 2025 17:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 24 05:02:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042325W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Apr 24 04:42:24 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	29398	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	178040	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	155882	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	86700	30.589	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.967%
60) 4-Bromofluorobenzene	12.847	95	83216	27.162	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	90.533%
63) Toluene-d8	10.565	98	259313	30.238	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042325\
Data File : VN086387.D
Acq On : 23 Apr 2025 17:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 24 05:02:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042325W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Apr 24 04:42:24 2025
Response via : Initial Calibration

