

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042419\
 Data File : VN055219.D
 Acq On : 24 Apr 2019 11:29
 Operator : JC/SP
 Sample : PB119160TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB119160TB

Quant Time: Apr 25 04:37:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	381449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	655759	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	594978	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	180163	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	264476	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
35) Dibromofluoromethane	7.59	113	214867	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
50) Toluene-d8	10.09	98	808258	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
62) 4-Bromofluorobenzene	12.40	95	251671	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
Target Compounds						
20) Methylene Chloride	4.55	84	27813	5.668	ug/l	97
80) 1,3,5-Trimethylbenzene	12.73	105	10001	0.742	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042419\
Data File : VN055219.D
Acq On : 24 Apr 2019 11:29
Operator : JC/SP
Sample : PB119160TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB119160TB

Quant Time: Apr 25 04:37:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

