

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042419\
 Data File : VN055220.D
 Acq On : 24 Apr 2019 11:55
 Operator : JC/SP
 Sample : PB119159TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB119159TB

Quant Time: Apr 25 04:39:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	375835	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	655444	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	593834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	176447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	257833	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
35) Dibromofluoromethane	7.59	113	210679	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
50) Toluene-d8	10.09	98	802771	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
62) 4-Bromofluorobenzene	12.40	95	250055	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
Target Compounds						
20) Methylene Chloride	4.55	84	29004	5.999	ug/l	96
43) Isopropyl Acetate	8.17	43	19886	2.501	ug/l #	86
80) 1,3,5-Trimethylbenzene	12.73	105	10004	0.757	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042419\
Data File : VN055220.D
Acq On : 24 Apr 2019 11:55
Operator : JC/SP
Sample : PB119159TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB119159TB

Quant Time: Apr 25 04:39:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

