

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042419\
 Data File : VN055225.D
 Acq On : 24 Apr 2019 14:02
 Operator : JC/SP
 Sample : K2542-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Apr 25 05:18:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	371208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	658447	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	596139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	183011	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	260178	51.78	ug/l	0.00
Spiked Amount						
			Recovery	=	103.56%	
35) Dibromofluoromethane	7.59	113	209405	46.81	ug/l	0.00
Spiked Amount						
			Recovery	=	93.62%	
50) Toluene-d8	10.09	98	803300	48.47	ug/l	0.00
Spiked Amount						
			Recovery	=	96.94%	
62) 4-Bromofluorobenzene	12.40	95	248506	47.35	ug/l	0.00
Spiked Amount						
			Recovery	=	94.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	10885	6.553	ug/l	96
18) Methyl Acetate	4.34	43	6789	0.793	ug/l	96
20) Methylene Chloride	4.55	84	125228	26.226	ug/l	97
30) Chloroform	7.37	83	4814	0.625	ug/l	95
80) 1,3,5-Trimethylbenzene	12.73	105	11086	0.809	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	22504	1.741	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042419\
Data File : VN055225.D
Acq On : 24 Apr 2019 14:02
Operator : JC/SP
Sample : K2542-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Apr 25 05:18:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

