


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042419\
Data File : VN055214.D
Acq On : 24 Apr 2019 8:35
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Title : SW846 8260
Last Update : Tue Apr 23 05:52:53 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	15355	PASS
75	95	30	60	52.8	37746	PASS
95	95	100	100	100.0	71522	PASS
96	95	5	9	6.5	4681	PASS
173	174	0.00	2	1.4	836	PASS
174	95	50	100	86.6	61906	PASS
175	174	5	9	6.5	4025	PASS
176	174	95	101	98.5	60952	PASS
177	176	5	9	6.3	3861	PASS