

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042423\  
 Data File : VN077438.D  
 Acq On : 24 Apr 2023 14:03  
 Operator : JC\MD  
 Sample : IBLK  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 IBLK

Quant Time: Apr 25 02:05:11 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042423W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 25 01:54:53 2023  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 8.236  | 168  | 583668   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 9.107  | 114  | 1091367  | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.866 | 117  | 967233   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.795 | 152  | 349828   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.583  | 65    | 378615   | 49.739   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 99.480%  |
| 35) Dibromofluoromethane    | 8.172  | 113   | 336958   | 51.512   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 103.020% |
| 50) Toluene-d8              | 10.572 | 98    | 1327240  | 53.057   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 106.120% |
| 62) 4-Bromofluorobenzene    | 12.854 | 95    | 409137   | 49.898   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 99.800%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042423\  
Data File : VN077438.D  
Acq On : 24 Apr 2023 14:03  
Operator : JC\MD  
Sample : IBLK  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
IBLK

Quant Time: Apr 25 02:05:11 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042423W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 25 01:54:53 2023  
Response via : Initial Calibration

