

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077445.D
 Acq On : 25 Apr 2023 04:04
 Operator : JC\MD
 Sample : VN0424WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0424WBL01

Quant Time: Apr 25 08:38:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	508665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	945593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	833763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	285744	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	332882	50.152	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.300%
35) Dibromofluoromethane	8.172	113	296329	52.139	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.280%
50) Toluene-d8	10.572	98	1147858	52.697	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.400%
62) 4-Bromofluorobenzene	12.854	95	348534	48.906	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077445.D
Acq On : 25 Apr 2023 04:04
Operator : JC\MD
Sample : VN0424WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0424WBL01

Quant Time: Apr 25 08:38:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

