

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077449.D
 Acq On : 25 Apr 2023 05:40
 Operator : JC\MD
 Sample : PB152304TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152304TB

Quant Time: Apr 25 08:41:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	555861	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1033522	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	947370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	327434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	366775	50.566	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.140%	
35) Dibromofluoromethane	8.172	113	311974	50.222	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.440%	
50) Toluene-d8	10.572	98	1273011	53.471	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.940%	
62) 4-Bromofluorobenzene	12.854	95	407133	52.268	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.540%	
Target Compounds						
						Qvalue
16) Acetone	4.443	43	39271	17.018	ug/l	88
20) Methylene Chloride	5.290	84	45116	6.165	ug/l	94
37) Ethyl Acetate	7.566	43	102614	15.047	ug/l	100
43) Isopropyl Acetate	8.689	43	429818	36.432	ug/l #	90
93) 1,2,4-Trichlorobenzene	15.365	180	2758	1.946	ug/l	93
95) Naphthalene	15.589	128	11818	2.468	ug/l	98
96) 1,2,3-Trichlorobenzene	15.765	180	3491	2.058	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077449.D
Acq On : 25 Apr 2023 05:40
Operator : JC\MD
Sample : PB152304TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152304TB

Quant Time: Apr 25 08:41:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

