

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077452.D
 Acq On : 25 Apr 2023 06:53
 Operator : JC\MD
 Sample : PB152304ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152304ZHE#03

Quant Time: Apr 25 08:42:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	571478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1086568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1017183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	369974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	382197	51.253	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.500%			
35) Dibromofluoromethane	8.172	113	327814	50.196	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.400%			
50) Toluene-d8	10.572	98	1342001	53.616	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.240%			
62) 4-Bromofluorobenzene	12.848	95	432600	52.826	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 105.660%			
Target Compounds					Qvalue	
16) Acetone	4.443	43	66442	28.006	ug/l	94
20) Methylene Chloride	5.284	84	39444	5.243	ug/l #	92
37) Ethyl Acetate	7.572	43	103927	14.496	ug/l	96
43) Isopropyl Acetate	8.695	43	521318	42.031	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077452.D
Acq On : 25 Apr 2023 06:53
Operator : JC\MD
Sample : PB152304ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152304ZHE#03

Quant Time: Apr 25 08:42:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

