

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077456.D
 Acq On : 25 Apr 2023 08:30
 Operator : JC\MD
 Sample : 02464-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-91

Quant Time: Apr 25 09:12:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	588899	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1110390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1015776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	362734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	377512	49.127	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.260%
35) Dibromofluoromethane	8.172	113	332382	49.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.600%
50) Toluene-d8	10.572	98	1361174	53.216	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.440%
62) 4-Bromofluorobenzene	12.848	95	435640	52.056	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.120%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	73404	30.025	ug/l	97
20) Methylene Chloride	5.290	84	35503	4.579	ug/l	94
37) Ethyl Acetate	7.572	43	97701	13.335	ug/l	97
43) Isopropyl Acetate	8.695	43	518473	40.904	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077456.D
Acq On : 25 Apr 2023 08:30
Operator : JC\MD
Sample : 02464-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-91

Quant Time: Apr 25 09:12:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

