

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077459.D
 Acq On : 25 Apr 2023 09:42
 Operator : JC\MD
 Sample : 02423-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPD-2WC-23-04-18

Quant Time: Apr 25 14:06:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	561510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1041984	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	971069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	341630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	373765	51.012	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.020%		
35) Dibromofluoromethane	8.172	113	326900	52.197	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.400%		
50) Toluene-d8	10.572	98	1308303	54.507	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 109.020%		
62) 4-Bromofluorobenzene	12.854	95	405313	51.612	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.220%		
Target Compounds						
16) Acetone	4.442	43	71923	30.855	ug/l	93
20) Methylene Chloride	5.289	84	30870	4.176	ug/l	92
37) Ethyl Acetate	7.572	43	70678	10.280	ug/l	99
43) Isopropyl Acetate	8.695	43	394383	33.157	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077459.D
Acq On : 25 Apr 2023 09:42
Operator : JC\MD
Sample : 02423-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPD-2WC-23-04-18

Quant Time: Apr 25 14:06:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

