

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077460.D
 Acq On : 25 Apr 2023 10:06
 Operator : JC\MD
 Sample : 02423-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPE-1WC-23-04-18

Quant Time: Apr 25 14:07:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	685648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1267426	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1158830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	411152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	448211	50.097	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.200%	
35) Dibromofluoromethane	8.172	113	386636	50.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.500%	
50) Toluene-d8	10.566	98	1571431	53.824	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.640%	
62) 4-Bromofluorobenzene	12.854	95	481220	50.378	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	145242	51.027	ug/l	97
20) Methylene Chloride	5.284	84	31904	3.534	ug/l #	88
37) Ethyl Acetate	7.566	43	87882	10.509	ug/l	98
43) Isopropyl Acetate	8.689	43	456751	31.570	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077460.D
Acq On : 25 Apr 2023 10:06
Operator : JC\MD
Sample : 02423-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPE-1WC-23-04-18

Quant Time: Apr 25 14:07:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

