

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077435.D
 Acq On : 24 Apr 2023 12:51
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Quant Time: Apr 25 02:03:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 01:54:53 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 04/25/2023
 Supervised By : Mahesh Dadoda 04/25/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	649119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1222453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	1083716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	439532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	168615	19.918	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	39.840%#		
35) Dibromofluoromethane	8.172	113	145368	19.840	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	39.680%#		
50) Toluene-d8	10.572	98	568011	20.272	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	40.540%#		
62) 4-Bromofluorobenzene	12.854	95	185332	20.179	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	40.360%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.137	85	162032	19.743	ug/l	100
3) Chloromethane	2.372	50	126006	19.365	ug/l	98
4) Vinyl Chloride	2.525	62	171317	19.764	ug/l	98
5) Bromomethane	2.960	94	140611	17.135	ug/l	92
6) Chloroethane	3.131	64	123106	18.694	ug/l	95
7) Trichlorofluoromethane	3.507	101	253943	19.932	ug/l	95
8) Diethyl Ether	3.978	74	91582	19.782	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.390	101	141185	19.055	ug/l	98
10) Methyl Iodide	4.602	142	177374	19.641	ug/l	97
11) Tert butyl alcohol	5.525	59	131448	105.131	ug/l #	88
12) 1,1-Dichloroethene	4.349	96	138583	19.029	ug/l	94
13) Acrolein	4.190	56	130292	97.229	ug/l	94
14) Allyl chloride	5.037	41	163405	20.289	ug/l	99
15) Acrylonitrile	5.731	53	301557	97.758	ug/l	99
16) Acetone	4.437	43	242025	89.814	ug/l	97
17) Carbon Disulfide	4.725	76	349861	19.007	ug/l	98
18) Methyl Acetate	5.037	43	200386	19.256	ug/l	98
19) Methyl tert-butyl Ether	5.807	73	472475	19.612	ug/l	99
20) Methylene Chloride	5.284	84	155225	18.164	ug/l	97
21) trans-1,2-Dichloroethene	5.790	96	146694	18.727	ug/l	92
22) Diisopropyl ether	6.684	45	361509	19.558	ug/l	97
23) Vinyl Acetate	6.613	43	1115745	96.655	ug/l	96
24) 1,1-Dichloroethane	6.578	63	249684	19.083	ug/l	98
25) 2-Butanone	7.490	43	367607	96.181	ug/l	98
26) 2,2-Dichloropropane	7.495	77	242113	19.084	ug/l	99
27) cis-1,2-Dichloroethene	7.490	96	178574	19.249	ug/l	98
28) Bromochloromethane	7.813	49	103454	20.705	ug/l #	5
29) Tetrahydrofuran	7.842	42	237397	99.134	ug/l	97
30) Chloroform	7.972	83	288295	19.551	ug/l	92
31) Cyclohexane	8.260	56	225319	15.938	ug/l	96
32) 1,1,1-Trichloroethane	8.172	97	269352	19.869	ug/l	98
36) 1,1-Dichloropropene	8.378	75	210637	19.843	ug/l	97
37) Ethyl Acetate	7.566	43	146798	18.199	ug/l	98
38) Carbon Tetrachloride	8.372	117	228356	19.498	ug/l	98
39) Methylcyclohexane	9.601	83	270722	19.357	ug/l	96
40) Benzene	8.613	78	620943	19.202	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077435.D
 Acq On : 24 Apr 2023 12:51
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/25/2023
 Supervised By : Mahesh Dadoda 04/25/2023

Quant Time: Apr 25 02:03:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 01:54:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.784	41	81668	19.439	ug/l	96
42) 1,2-Dichloroethane	8.672	62	210437	18.947	ug/l	99
43) Isopropyl Acetate	8.695	43	267721	19.185	ug/l	98
44) Trichloroethene	9.354	130	159770	19.297	ug/l	100
45) 1,2-Dichloropropane	9.625	63	143975	19.007	ug/l	97
46) Dibromomethane	9.707	93	114798	18.677	ug/l	97
47) Bromodichloromethane	9.889	83	228546	19.223	ug/l	98
48) Methyl methacrylate	9.684	41	121718	19.162	ug/l	99
49) 1,4-Dioxane	9.695	88	59707	374.742	ug/l	96
51) 4-Methyl-2-Pentanone	10.448	43	791194	96.665	ug/l	99
52) Toluene	10.631	92	424866	20.056	ug/l	98
53) t-1,3-Dichloropropene	10.836	75	233562	19.401	ug/l	98
54) cis-1,3-Dichloropropene	10.313	75	254893	19.452	ug/l	100
55) 1,1,2-Trichloroethane	11.019	97	161916	19.007	ug/l	97
56) Ethyl methacrylate	10.878	69	233677	19.576	ug/l	97
57) 1,3-Dichloropropane	11.166	76	265343	19.555	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.166	63	364187	90.203	ug/l	99
59) 2-Hexanone	11.195	43	569753	95.764	ug/l	100
60) Dibromochloromethane	11.360	129	169483	19.170	ug/l	99
61) 1,2-Dibromoethane	11.472	107	161794	18.603	ug/l	100
64) Tetrachloroethene	11.107	164	123142	18.982	ug/l	97
65) Chlorobenzene	11.895	112	439622	19.170	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.966	131	162543	19.482	ug/l	99
67) Ethyl Benzene	11.966	91	796867	19.739	ug/l	99
68) m/p-Xylenes	12.072	106	613100	39.076	ug/l	99
69) o-Xylene	12.401	106	311925	20.212	ug/l	97
70) Styrene	12.413	104	482026	19.972	ug/l	99
71) Bromoform	12.583	173	104088	19.069	ug/l #	99
73) Isopropylbenzene	12.695	105	777195	19.055	ug/l	99
74) N-amyl acetate	12.495	43	225150	18.646	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.942	83	236838	18.481	ug/l	98
76) 1,2,3-Trichloropropane	12.995	75	195688m	18.737	ug/l	
77) Bromobenzene	12.983	156	166781	18.762	ug/l	99
78) n-propylbenzene	13.036	91	870923	19.312	ug/l	100
79) 2-Chlorotoluene	13.130	91	536570	19.298	ug/l	98
80) 1,3,5-Trimethylbenzene	13.177	105	608862	20.101	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.736	75	68409	19.443	ug/l	96
82) 4-Chlorotoluene	13.224	91	492479	18.955	ug/l	99
83) tert-Butylbenzene	13.442	119	560874	19.068	ug/l	99
84) 1,2,4-Trimethylbenzene	13.483	105	568313	20.090	ug/l	99
85) sec-Butylbenzene	13.619	105	775621	19.211	ug/l	99
86) p-Isopropyltoluene	13.730	119	604020	19.444	ug/l	99
87) 1,3-Dichlorobenzene	13.736	146	302197	19.116	ug/l	97
88) 1,4-Dichlorobenzene	13.813	146	289207	18.543	ug/l	98
89) n-Butylbenzene	14.060	91	466967	19.424	ug/l	100
90) Hexachloroethane	14.336	117	126046	18.366	ug/l	97
91) 1,2-Dichlorobenzene	14.107	146	298295	19.025	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.724	75	43763	18.467	ug/l	99
93) 1,2,4-Trichlorobenzene	15.366	180	68938	18.569	ug/l	97
94) Hexachlorobutadiene	15.466	225	59281	18.710	ug/l	98
95) Naphthalene	15.589	128	255230	18.670	ug/l	99
96) 1,2,3-Trichlorobenzene	15.754	180	64782	18.581	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077435.D
Acq On : 24 Apr 2023 12:51
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023

Quant Time: Apr 25 02:03:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 01:54:53 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077435.D
 Acq On : 24 Apr 2023 12:51
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Manual Integrations APPROVED

Reviewed By : John Carlone 04/25/2023
 Supervised By : Mahesh Dadoda 04/25/2023


Quant Time: Apr 25 02:03:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 01:54:53 2023
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.231 min Scan# 1070
Delta R.T. -0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:168 Resp: 649119
Ion Ratio Lower Upper
168 100
99 62.8 53.0 79.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#2
Dichlorodifluoromethane
Concen: 19.743 ug/l
RT: 2.137 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion: 85 Resp: 162032
Ion Ratio Lower Upper
85 100
87 31.4 15.8 47.4





#3
Chloromethane
Concen: 19.365 ug/l
RT: 2.372 min Scan# 74
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 50 Resp: 126000
Ion Ratio Lower Upper
50 100
52 34.2 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#4
Vinyl Chloride
Concen: 19.764 ug/l
RT: 2.525 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Tgt Ion: 62 Resp: 171317
Ion Ratio Lower Upper
62 100
64 31.6 26.2 39.2





#5
Bromomethane
Concen: 17.135 ug/l
RT: 2.960 min Scan# 174
Delta R.T. 0.012 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 94 Resp: 14061
Ion Ratio Lower Upper
94 100
96 98.7 73.1 109.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#6
Chloroethane
Concen: 18.694 ug/l
RT: 3.131 min Scan# 203
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion: 64 Resp: 123106
Ion Ratio Lower Upper
64 100
66 34.8 25.4 38.2





#7
 Trichlorofluoromethane
 Concen: 19.932 ug/l
 RT: 3.507 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020



Tgt Ion:101 Resp: 25394
 Ion Ratio Lower Upper
 101 100
 103 59.4 50.7 76.1




Manual Integrations
 APPROVED


Reviewed By :John Carlone 04/25/2023
 Supervised By :Mahesh Dadoda 04/25/2023



#8
 Diethyl Ether
 Concen: 19.782 ug/l
 RT: 3.978 min Scan# 347
 Delta R.T. 0.006 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51

Tgt Ion: 74 Resp: 91582
 Ion Ratio Lower Upper
 74 100
 45 68.0 34.4 103.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.055 ug/l
RT: 4.390 min Scan# 415
Delta R.T. 0.012 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:101 Resp: 14118
Ion Ratio Lower Upper
101 100
85 45.3 36.6 55.0
151 74.5 57.6 86.4




Manual Integrations
APPROVED


Reviewed By :John Carbone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#10
Methyl Iodide
Concen: 19.641 ug/l
RT: 4.602 min Scan# 453
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion:142 Resp: 177374
Ion Ratio Lower Upper
142 100
127 45.0 37.6 56.4
141 15.4 11.4 17.0





#11
Tert butyl alcohol
Concen: 105.131 ug/l
RT: 5.525 min Scan# 610
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
Client Sample Id :
VSTDIC020




Tgt Ion: 59 Resp: 131448
Ion Ratio Lower Upper
59 100
57 5.8 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#12
1,1-Dichloroethene
Concen: 19.029 ug/l
RT: 4.349 min Scan# 410
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51



Tgt Ion: 96 Resp: 138583
Ion Ratio Lower Upper
96 100
61 132.8 112.7 169.1
98 61.3 51.2 76.8









#13
Acrolein
Concen: 97.229 ug/l
RT: 4.190 min Scan# 383
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 56 Resp: 130292
Ion Ratio Lower Upper
56 100
55 72.8 54.5 81.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#14
Allyl chloride
Concen: 20.289 ug/l
RT: 5.037 min Scan# 527
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion: 41 Resp: 163405
Ion Ratio Lower Upper
41 100
39 73.7 58.8 88.2
76 51.2 39.1 58.7





#15
Acrylonitrile
Concen: 97.758 ug/l
RT: 5.731 min Scan# 645
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion: 53 Resp: 30155
Ion Ratio Lower Upper
53 100
52 82.6 65.3 97.9
51 35.9 28.9 43.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#16
Acetone
Concen: 89.814 ug/l
RT: 4.437 min Scan# 425
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Tgt Ion: 43 Resp: 242025
Ion Ratio Lower Upper
43 100
58 36.8 28.0 42.0





#17
Carbon Disulfide
Concen: 19.007 ug/l
RT: 4.725 min Scan# 474
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion: 76 Resp: 34986
Ion Ratio Lower Upper
76 100
78 8.6 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#18
Methyl Acetate
Concen: 19.256 ug/l
RT: 5.037 min Scan# 527
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Tgt Ion: 43 Resp: 200386
Ion Ratio Lower Upper
43 100
74 31.7 26.2 39.4





#19
Methyl tert-butyl Ether
Concen: 19.612 ug/l
RT: 5.807 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 73 Resp: 472475
Ion Ratio Lower Upper
73 100
57 19.8 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#20
Methylene Chloride
Concen: 18.164 ug/l
RT: 5.284 min Scan# 569
Delta R.T. -0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion: 84 Resp: 155225
Ion Ratio Lower Upper
84 100
49 102.8 79.7 119.5
51 30.4 25.2 37.8
86 68.9 53.3 79.9





#21
trans-1,2-Dichloroethene
Concen: 18.727 ug/l
RT: 5.790 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 96 Resp: 146694
Ion Ratio Lower Upper
96 100
61 119.3 90.1 135.1
98 65.5 46.4 69.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#22
Diisopropyl ether
Concen: 19.558 ug/l
RT: 6.684 min Scan# 807
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion: 45 Resp: 361509
Ion Ratio Lower Upper
45 100
43 54.1 45.5 68.3
87 39.2 31.8 47.8
59 16.1 11.6 17.4









#23
 Vinyl Acetate
 Concen: 96.655 ug/l
 RT: 6.613 min Scan# 795
 Delta R.T. 0.000 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51

Instrument :
 MSVOA_N
 Client Sample Id :
 VSTDIC020

Tgt Ion: 43 Resp: 111574
 Ion Ratio Lower Upper
 43 100
 86 14.7 13.0 19.6




Manual Integrations
 APPROVED


Reviewed By :John Carlone 04/25/2023
 Supervised By :Mahesh Dadoda 04/25/2023



#24
 1,1-Dichloroethane
 Concen: 19.083 ug/l
 RT: 6.578 min Scan# 789
 Delta R.T. 0.000 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51

Tgt Ion: 63 Resp: 249684
 Ion Ratio Lower Upper
 63 100
 98 8.4 3.9 11.7
 100 5.8 2.5 7.4









#25
2-Butanone
Concen: 96.181 ug/l
RT: 7.490 min Scan# 944
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 43 Resp: 36760
Ion Ratio Lower Upper
43 100
72 36.3 30.2 45.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#26
2,2-Dichloropropane
Concen: 19.084 ug/l
RT: 7.495 min Scan# 945
Delta R.T. -0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion: 77 Resp: 242113
Ion Ratio Lower Upper
77 100
97 22.9 11.2 33.6





#27

cis-1,2-Dichloroethene

Concen: 19.249 ug/l

RT: 7.490 min Scan# 94

Delta R.T. -0.006 min

Lab File: VN077435.D


Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020



Tgt Ion: 96 Resp: 178574

Ion Ratio Lower Upper

96 100

61 124.6 0.0 255.2




98 64.9 0.0 128.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#28

Bromochloromethane

Concen: 20.705 ug/l

RT: 7.813 min Scan# 999

Delta R.T. -0.006 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51




Tgt Ion: 49 Resp: 103454


Ion Ratio Lower Upper

49 100

129 1.0 0.0 4.8

130 0.0 75.3 112.9#





#29

Tetrahydrofuran

Concen: 99.134 ug/l

RT: 7.842 min Scan# 1004

Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 42 Resp: 237397
 Ion Ratio Lower Upper
 42 100
 72 61.1 49.8 74.6
 71 56.6 47.4 71.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#30

Chloroform

Concen: 19.551 ug/l




RT: 7.972 min Scan# 1026


Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51


Tgt Ion: 83 Resp: 288295
 Ion Ratio Lower Upper
 83 100
 85 63.0 55.4 83.0





#31
Cyclohexane
Concen: 15.938 ug/l
RT: 8.260 min Scan# 1075
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 56 Resp: 225319
Ion Ratio Lower Upper
56 100
69 37.4 26.8 40.2
84 104.6 81.8 122.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#32
1,1,1-Trichloroethane
Concen: 19.869 ug/l
RT: 8.172 min Scan# 1060
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion: 97 Resp: 269352
Ion Ratio Lower Upper
97 100
99 60.4 50.3 75.5
61 42.6 34.7 52.1





#33

1,2-Dichloroethane-d4

Concen: 19.918 ug/l

RT: 8.584 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 16861

Ion Ratio Lower Upper

65 100





67 51.2 0.0 106.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.107 min Scan# 1219

Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51




Tgt Ion:114 Resp: 1222453


Ion Ratio Lower Upper

114 100

63 19.1 0.0 36.4

88 15.1 0.0 30.6





#35

Dibromofluoromethane

Concen: 19.840 ug/l

RT: 8.172 min Scan# 1060

Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 113 Resp: 14536

Ion Ratio Lower Upper

113 100

111 101.6 82.6 123.8





192 16.6 12.7 19.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#36

1,1-Dichloropropene

Concen: 19.843 ug/l

RT: 8.378 min Scan# 1095

Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51




Tgt Ion: 75 Resp: 210637


Ion Ratio Lower Upper

75 100

110 35.7 19.1 57.3

77 30.6 25.0 37.4





#37

Ethyl Acetate

Concen: 18.199 ug/l

RT: 7.566 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 146798





Ion	Ratio	Lower	Upper
43	100		
61	17.0	13.2	19.8
70	13.1	11.3	16.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#38

Carbon Tetrachloride

Concen: 19.498 ug/l

RT: 8.372 min Scan# 1094




Delta R.T. 0.006 min


Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Tgt Ion:117 Resp: 228356

Ion	Ratio	Lower	Upper
117	100		
119	98.8	77.3	115.9
121	30.1	23.3	34.9





#39

Methylcyclohexane

Concen: 19.357 ug/l

RT: 9.601 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN077435.D


Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	83	Resp:	27072
Ion Ratio	100	Lower	Upper
83	100		
55	65.0	49.1	73.7
98	49.0	40.6	61.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#40

Benzene

Concen: 19.202 ug/l




RT: 8.613 min Scan# 1135


Delta R.T. 0.006 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51


Tgt Ion:	78	Resp:	620943
Ion Ratio	100	Lower	Upper
78	100		
77	23.1	19.0	28.6





#41
Methacrylonitrile
Concen: 19.439 ug/l
RT: 7.784 min Scan# 994
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion: 41 Resp: 81668
Ion Ratio Lower Upper
41 100
39 58.7 42.6 64.0
67 95.9 77.5 116.3
52 40.0 28.9 43.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#42
1,2-Dichloroethane
Concen: 18.947 ug/l
RT: 8.672 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Tgt Ion: 62 Resp: 210437
Ion Ratio Lower Upper
62 100
98 10.1 0.0 19.8





#43
Isopropyl Acetate
Concen: 19.185 ug/l
RT: 8.695 min Scan# 1149
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 43 Resp: 26772
Ion Ratio Lower Upper
43 100
61 32.2 24.7 37.1
87 18.8 14.6 21.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#44
Trichloroethene
Concen: 19.297 ug/l
RT: 9.354 min Scan# 1261
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51



Tgt Ion:130 Resp: 159770
Ion Ratio Lower Upper
130 100
95 102.2 0.0 204.6





#45

1,2-Dichloropropane

Concen: 19.007 ug/l

RT: 9.625 min Scan# 1307

Delta R.T. 0.006 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 143978

Ion Ratio Lower Upper

63 100





65 31.8 23.9 35.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#46

Dibromomethane

Concen: 18.677 ug/l

RT: 9.707 min Scan# 1321

Delta R.T. -0.006 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51




Tgt Ion: 93 Resp: 114798


Ion Ratio Lower Upper

93 100

95 81.6 63.8 95.8


174 85.2 65.6 98.4





#47
Bromodichloromethane
Concen: 19.223 ug/l
RT: 9.889 min Scan# 1352
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 83 Resp: 228540
Ion Ratio Lower Upper
83 100
85 63.3 49.6 74.4
127 8.2 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#48
Methyl methacrylate
Concen: 19.162 ug/l
RT: 9.684 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51



Tgt Ion: 41 Resp: 121718
Ion Ratio Lower Upper
41 100
69 111.4 90.4 135.6
39 67.4 54.6 81.8





#49

1,4-Dioxane

Concen: 374.742 ug/l

RT: 9.695 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	88	Resp:	5970
Ion Ratio	Lower	Upper	
88	100		
43	21.1	17.0	25.6
58	62.2	47.0	70.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#50

Toluene-d8

Concen: 20.272 ug/l




RT: 10.572 min Scan# 1468


Delta R.T. 0.006 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51


Tgt Ion:	98	Resp:	568011
Ion Ratio	Lower	Upper	
98	100		
100	63.9	52.0	78.0





#51
4-Methyl-2-Pentanone
Concen: 96.665 ug/l
RT: 10.448 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 791194
Ion Ratio Lower Upper
43 100
58 46.1 36.6 54.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#52
Toluene
Concen: 20.056 ug/l
RT: 10.631 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion: 92 Resp: 424866
Ion Ratio Lower Upper
92 100
91 168.9 137.4 206.0





#53
 t-1,3-Dichloropropene
 Concen: 19.401 ug/l
 RT: 10.836 min Scan# 1513
 Delta R.T. 0.000 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020




Tgt Ion: 75 Resp: 233562
 Ion Ratio Lower Upper
 75 100
 77 32.7 25.4 38.0

Manual Integrations
 APPROVED



Reviewed By :John Carlone 04/25/2023
 Supervised By :Mahesh Dadoda 04/25/2023




#54
 cis-1,3-Dichloropropene
 Concen: 19.452 ug/l
 RT: 10.313 min Scan# 1424
 Delta R.T. 0.000 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51




Tgt Ion: 75 Resp: 254893
 Ion Ratio Lower Upper
 75 100
 77 32.4 26.1 39.1
 39 39.5 31.6 47.4





#55
1,1,2-Trichloroethane
Concen: 19.007 ug/l
RT: 11.019 min Scan# 1544
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 97 Resp: 161910
Ion Ratio Lower Upper
97 100
83 82.7 65.1 97.7
85 53.2 43.4 65.2
99 57.7 49.2 73.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#56
Ethyl methacrylate
Concen: 19.576 ug/l
RT: 10.878 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion: 69 Resp: 233677
Ion Ratio Lower Upper
69 100
41 50.2 39.6 59.4
39 32.5 28.6 43.0





#57
1,3-Dichloropropane
Concen: 19.555 ug/l
RT: 11.166 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 76 Resp: 26534
Ion Ratio Lower Upper
76 100
78 32.6 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#58
2-Chloroethyl Vinyl ether
Concen: 90.203 ug/l
RT: 10.166 min Scan# 1399
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion: 63 Resp: 364187
Ion Ratio Lower Upper
63 100
106 32.3 26.2 39.4





#59
2-Hexanone
Concen: 95.764 ug/l
RT: 11.195 min Scan# 1574
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 43 Resp: 56975
Ion Ratio Lower Upper
43 100
58 63.8 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#60
Dibromochloromethane
Concen: 19.170 ug/l
RT: 11.360 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion:129 Resp: 169483
Ion Ratio Lower Upper
129 100
127 78.1 38.6 116.0





#61
1,2-Dibromoethane
Concen: 18.603 ug/l
RT: 11.472 min Scan# 1621
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion:107 Resp: 161794
Ion Ratio Lower Upper
107 100
109 95.9 76.6 114.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#62
4-Bromofluorobenzene
Concen: 20.179 ug/l
RT: 12.854 min Scan# 1856
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Tgt Ion: 95 Resp: 185332
Ion Ratio Lower Upper
95 100
174 71.5 0.0 143.8
176 67.9 0.0 135.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.872 min Scan# 1688
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion:117 Resp: 1083710
Ion Ratio Lower Upper
117 100
82 54.4 44.0 66.0
119 31.5 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#64
Tetrachloroethene
Concen: 18.982 ug/l
RT: 11.107 min Scan# 1559
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion:164 Resp: 123142
Ion Ratio Lower Upper
164 100
166 131.4 101.4 152.0
129 105.8 82.4 123.6
131 98.7 78.2 117.4









#65
Chlorobenzene
Concen: 19.170 ug/l
RT: 11.895 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:112 Resp: 43962
Ion Ratio Lower Upper
112 100
114 32.3 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 19.482 ug/l
RT: 11.966 min Scan# 1705
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion:131 Resp: 162543
Ion Ratio Lower Upper
131 100
133 92.2 46.4 139.1
119 68.6 34.1 102.2





#67
Ethyl Benzene
Concen: 19.739 ug/l
RT: 11.966 min Scan# 1705
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 91 Resp: 79686
Ion Ratio Lower Upper
91 100
106 31.2 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#68
m/p-Xylenes
Concen: 39.076 ug/l
RT: 12.072 min Scan# 1723
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51



Tgt Ion:106 Resp: 613100
Ion Ratio Lower Upper
106 100
91 201.1 159.6 239.4









#69
o-Xylene
Concen: 20.212 ug/l
RT: 12.401 min Scan# 1779
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:106 Resp: 311925
Ion Ratio Lower Upper
106 100
91 206.1 105.2 315.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#70
Styrene
Concen: 19.972 ug/l
RT: 12.413 min Scan# 1781
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion:104 Resp: 482026
Ion Ratio Lower Upper
104 100
78 48.7 39.5 59.3
103 55.1 43.2 64.8





#71
Bromoform
Concen: 19.069 ug/l
RT: 12.583 min Scan# 1809
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion:173 Resp: 104088
Ion Ratio Lower Upper
173 100
175 49.3 24.3 73.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.795 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51


Tgt Ion:152 Resp: 439532
Ion Ratio Lower Upper
152 100
115 64.5 32.3 96.8
150 164.9 0.0 355.8





#73
Isopropylbenzene
Concen: 19.055 ug/l
RT: 12.695 min Scan# 1829
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion:105 Resp: 777195
Ion Ratio Lower Upper
105 100
120 26.8 13.1 39.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#74
N-ethyl acetate
Concen: 18.646 ug/l
RT: 12.495 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Tgt Ion: 43 Resp: 225150
Ion Ratio Lower Upper
43 100
70 55.0 43.4 65.2
55 28.3 21.4 32.0
61 28.9 22.8 34.2





#75
1,1,2,2-Tetrachloroethane
Concen: 18.481 ug/l
RT: 12.942 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 83 Resp: 236838
Ion Ratio Lower Upper
83 100
131 9.9 5.0 15.0
85 62.0 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#76
1,2,3-Trichloropropane
Concen: 18.737 ug/l m
RT: 12.995 min Scan# 1880
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Tgt Ion: 75 Resp: 195688
Ion Ratio Lower Upper
75 100
77 183.2 84.5 253.5





#77
Bromobenzene
Concen: 18.762 ug/l
RT: 12.983 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion:156 Resp: 16678
Ion Ratio Lower Upper
156 100
77 215.0 107.7 323.3
158 94.1 48.4 145.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#78
n-propylbenzene
Concen: 19.312 ug/l
RT: 13.036 min Scan# 1887
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51



Tgt Ion: 91 Resp: 870923
Ion Ratio Lower Upper
91 100
120 23.1 11.5 34.4









#79
2-Chlorotoluene
Concen: 19.298 ug/l
RT: 13.130 min Scan# 1902
Delta R.T. 0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 536570
Ion Ratio Lower Upper
91 100
126 33.6 16.3 48.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#80
1,3,5-Trimethylbenzene
Concen: 20.101 ug/l
RT: 13.177 min Scan# 1911
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion:105 Resp: 608862
Ion Ratio Lower Upper
105 100
120 49.1 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 19.443 ug/l
RT: 12.736 min Scan# 1836
Delta R.T. -0.006 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 75 Resp: 68409
Ion Ratio Lower Upper
75 100
53 89.7 67.4 101.2
89 49.2 38.8 58.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#82
4-Chlorotoluene
Concen: 18.955 ug/l
RT: 13.224 min Scan# 1919
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion: 91 Resp: 492479
Ion Ratio Lower Upper
91 100
126 33.9 16.7 50.0









#83
tert-Butylbenzene
Concen: 19.068 ug/l
RT: 13.442 min Scan# 1956
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:119 Resp: 560874
Ion Ratio Lower Upper
119 100
91 67.3 33.9 101.7
134 26.2 13.2 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#84
1,2,4-Trimethylbenzene
Concen: 20.090 ug/l
RT: 13.483 min Scan# 1963
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion:105 Resp: 568313
Ion Ratio Lower Upper
105 100
120 44.8 22.0 66.0









#85
 sec-Butylbenzene
 Concen: 19.211 ug/l
 RT: 13.619 min Scan# 1986
 Delta R.T. 0.000 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Tgt Ion:105 Resp: 77562
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.8 29.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 04/25/2023
 Supervised By :Mahesh Dadoda 04/25/2023



#86
 p-Isopropyltoluene
 Concen: 19.444 ug/l
 RT: 13.730 min Scan# 2005
 Delta R.T. 0.000 min
 Lab File: VN077435.D
 Acq: 24 Apr 2023 12:51


Tgt Ion:119 Resp: 604020
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.6 37.8
 91 24.6 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 19.116 ug/l
RT: 13.736 min Scan# 2006
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion:146 Resp: 30219
Ion Ratio Lower Upper
146 100
111 41.5 22.1 66.3
148 62.3 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023




#88
1,4-Dichlorobenzene
Concen: 18.543 ug/l
RT: 13.813 min Scan# 2019
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51



Tgt Ion:146 Resp: 289207
Ion Ratio Lower Upper
146 100
111 44.5 20.6 61.8
148 63.5 31.6 94.8









#89
n-Butylbenzene
Concen: 19.424 ug/l
RT: 14.060 min Scan# 2061
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 46696
Ion Ratio Lower Upper
91 100
92 52.4 26.4 79.0
134 25.1 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#90
Hexachloroethane
Concen: 18.366 ug/l
RT: 14.336 min Scan# 2108
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion:117 Resp: 126046
Ion Ratio Lower Upper
117 100
201 56.6 29.3 88.0









#91
1,2-Dichlorobenzene
Concen: 19.025 ug/l
RT: 14.107 min Scan# 2069
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:146 Resp: 298295
Ion Ratio Lower Upper
146 100
111 43.7 22.0 66.0
148 64.0 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.467 ug/l
RT: 14.724 min Scan# 2174
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion: 75 Resp: 43763
Ion Ratio Lower Upper
75 100
155 78.0 39.1 117.2
157 101.5 50.3 150.9





#93

1,2,4-Trichlorobenzene

Concen: 18.569 ug/l

RT: 15.366 min Scan# 21

Delta R.T. -0.005 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:180 Resp: 68938

Ion Ratio Lower Upper

180 100

182 88.8 46.1 138.2





145 31.8 16.6 49.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/25/2023

Supervised By :Mahesh Dadoda 04/25/2023



#94

Hexachlorobutadiene

Concen: 18.710 ug/l

RT: 15.466 min Scan# 2300

Delta R.T. 0.001 min

Lab File: VN077435.D

Acq: 24 Apr 2023 12:51




Tgt Ion:225 Resp: 59281


Ion Ratio Lower Upper

225 100

223 60.8 31.3 93.8


227 63.2 31.9 95.9





#95
Naphthalene
Concen: 18.670 ug/l
RT: 15.589 min Scan# 2321
Delta R.T. 0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion:128 Resp: 255230
Ion Ratio Lower Upper
128 100
127 12.8 10.0 15.0
129 10.4 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/25/2023
Supervised By :Mahesh Dadoda 04/25/2023



#96
1,2,3-Trichlorobenzene
Concen: 18.581 ug/l
RT: 15.754 min Scan# 2349
Delta R.T. -0.000 min
Lab File: VN077435.D
Acq: 24 Apr 2023 12:51

Tgt Ion:180 Resp: 64782
Ion Ratio Lower Upper
180 100
182 95.3 46.2 138.6
145 36.4 17.5 52.6

