

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077438.D
 Acq On : 24 Apr 2023 14:03
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 25 02:05:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 01:54:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	583668	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1091367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	967233	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	349828	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	378615	49.739	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.480%
35) Dibromofluoromethane	8.172	113	336958	51.512	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.020%
50) Toluene-d8	10.572	98	1327240	53.057	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.120%
62) 4-Bromofluorobenzene	12.854	95	409137	49.898	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.800%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077438.D
Acq On : 24 Apr 2023 14:03
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 25 02:05:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 01:54:53 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.236 min Scan# 1070
Delta R.T. 0.005 min
Lab File: VN077438.D
Acq: 24 Apr 2023 14:03




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 583668
Ion Ratio Lower Upper
168 100
99 63.2 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 49.739 ug/l
RT: 8.583 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN077438.D
Acq: 24 Apr 2023 14:03

Tgt Ion: 65 Resp: 378615
Ion Ratio Lower Upper
65 100
67 52.7 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.107 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN077438.D

Acq: 24 Apr 2023 14:03

Instrument :

MSVOA_N

ClientSampleId :

IBLK





Tgt Ion:114 Resp: 1091367

Ion Ratio Lower Upper

114 100

63 18.0 0.0 36.4

88 14.3 0.0 30.6



#35

Dibromofluoromethane

Concen: 51.512 ug/l

RT: 8.172 min Scan# 1060

Delta R.T. -0.000 min

Lab File: VN077438.D

Acq: 24 Apr 2023 14:03




Tgt Ion:113 Resp: 336958


Ion Ratio Lower Upper

113 100

111 103.9 82.6 123.8

192 15.8 12.7 19.1









#50
Toluene-d8
Concen: 53.057 ug/l
RT: 10.572 min Scan# 1468
Delta R.T. 0.006 min
Lab File: VN077438.D
Acq: 24 Apr 2023 14:03




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 1327240
Ion Ratio Lower Upper
98 100
100 64.4 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 49.898 ug/l
RT: 12.854 min Scan# 1856
Delta R.T. 0.006 min
Lab File: VN077438.D
Acq: 24 Apr 2023 14:03

Tgt Ion: 95 Resp: 409137
Ion Ratio Lower Upper
95 100
174 70.7 0.0 143.8
176 67.8 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.866 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN077438.D
Acq: 24 Apr 2023 14:03

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 967233
Ion Ratio Lower Upper
117 100
82 55.2 44.0 66.0
119 31.2 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.795 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN077438.D
Acq: 24 Apr 2023 14:03

Tgt Ion:152 Resp: 349828
Ion Ratio Lower Upper
152 100
115 63.4 32.3 96.8
150 155.1 0.0 355.8

