

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
 Data File : VN077463.D
 Acq On : 25 Apr 2023 11:19
 Operator : JC\MD
 Sample : 02460-03 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Apr 25 14:07:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	580585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1068696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	967028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	344542	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	373902	49.354	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.700%
35) Dibromofluoromethane	8.172	113	332810	51.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.620%
50) Toluene-d8	10.572	98	1310000	53.213	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.420%
62) 4-Bromofluorobenzene	12.854	95	401950	49.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.800%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042423\
Data File : VN077463.D
Acq On : 25 Apr 2023 11:19
Operator : JC\MD
Sample : 02460-03 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Apr 25 14:07:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.231 min Scan# 1070
Delta R.T. -0.000 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19




Instrument :
MSVOA_N
ClientSampleId :
MW-2


Tgt Ion:168 Resp: 580585
Ion Ratio Lower Upper
168 100
99 65.4 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 49.354 ug/l
RT: 8.584 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19

Tgt Ion: 65 Resp: 373902
Ion Ratio Lower Upper
65 100
67 51.4 0.0 106.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.107 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19




Instrument :
MSVOA_N
ClientSampleId :
MW-2


Tgt Ion:114 Resp: 1068696
Ion Ratio Lower Upper
114 100
63 17.9 0.0 36.4
88 15.0 0.0 30.6



#35
Dibromofluoromethane
Concen: 51.813 ug/l
RT: 8.172 min Scan# 1060
Delta R.T. -0.000 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19

Tgt Ion:113 Resp: 332810
Ion Ratio Lower Upper
113 100
111 101.4 82.6 123.8
192 15.7 12.7 19.1









#50
Toluene-d8
Concen: 53.213 ug/l
RT: 10.572 min Scan# 1468
Delta R.T. 0.006 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19




Instrument :
MSVOA_N
ClientSampleId :
MW-2


Tgt Ion: 98 Resp: 1310000
Ion Ratio Lower Upper
98 100
100 64.9 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 49.904 ug/l
RT: 12.854 min Scan# 1856
Delta R.T. 0.006 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19

Tgt Ion: 95 Resp: 401950
Ion Ratio Lower Upper
95 100
174 73.1 0.0 143.8
176 68.4 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.866 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Tgt Ion:117 Resp: 967028
Ion Ratio Lower Upper
117 100
82 55.2 44.0 66.0
119 30.9 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.795 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN077463.D
Acq: 25 Apr 2023 11:19

Tgt Ion:152 Resp: 344542
Ion Ratio Lower Upper
152 100
115 62.3 32.3 96.8
150 156.6 0.0 355.8

