


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042425\
Data File : VN086391.D
Acq On : 24 Apr 2025 10:36
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042325W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Thu Apr 24 04:42:24 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	9367	PASS
75	95	30	60	51.5	23715	PASS
95	95	100	100	100.0	46080	PASS
96	95	5	9	6.5	2983	PASS
173	174	0.00	2	1.0	335	PASS
174	95	50	100	72.5	33419	PASS
175	174	5	9	7.4	2457	PASS
176	174	95	101	96.7	32304	PASS
177	176	5	9	6.7	2180	PASS