

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042519\
 Data File : VN055244.D
 Acq On : 25 Apr 2019 14:18
 Operator : JC/SP
 Sample : K2203-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-03-042419-B

Quant Time: Apr 26 03:27:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	398676	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	697082	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	638756	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	196215	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	274368	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
35) Dibromofluoromethane	7.59	113	224067	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	10.09	98	864729	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
62) 4-Bromofluorobenzene	12.40	95	277316	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
Target Compounds						
16) Acetone	3.82	43	12487	7.000	ug/l	96
20) Methylene Chloride	4.55	84	1071245	208.886	ug/l	99
43) Isopropyl Acetate	8.17	43	20199	2.389	ug/l #	85
95) Naphthalene	15.13	128	6700	4.381	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042519\
Data File : VN055244.D
Acq On : 25 Apr 2019 14:18
Operator : JC/SP
Sample : K2203-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-03-042419-B

Quant Time: Apr 26 03:27:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

